



Republic of the Philippines
PROVINCE OF DAVAO DEL SUR
Barangay Matti, Digos City, Davao del Sur

OFFICE OF THE BIDS AND AWARDS COMMITTEE

SUPPLEMENTAL BID BULLETIN


ADDENDUM NO. 21-2020

This is in reference to the following publication made last October 27, 2020 for IB No. G-177-20.

Correction of information is made as follows:

G-177-20 (PR 1003-CB-20) (Supply / Delivery of 2 units Mobile Medical Radiography X-Ray System)	
FROM:	TO:
Additional Specifications To Comply:	
ACCESSORIES	
<ul style="list-style-type: none">- Radiologist Workstation (with Mini PACAS Functions)<ul style="list-style-type: none">• (Latest 7th generation Intel Core i5 Processor with Windows 10 operating software; minimum of 8GB RAM; minimum 1TB HDD; 21-inch or bigger Medical Grade LED Monitor)• Supported by all necessary software for all various Digital Radiography functions• The viewing stations must be provided with appropriate On-line Uninterrupted Power Supply (UPS)• Provision for all accessories like mouse, keyboard, power cable- One (1) pc. Lead apron with minimum lead equivalence of 0.25 mm lead- One (1) pc. Thyroid shield- One (1) pc. measuring calliper, sliding, double sided, scaled in cm, with blunt edges and parallel arm Abdominal shield (for pregnant women)	
Warranty: 5 years warranty (unlimited service) free parts and labor Should Have a service center within Region XI	Warranty: 1.) 5 years warranty for unlimited service which includes but not limited to, preventive maintenance services 2.) 3 years warranty for parts and labor 3.) Should have a service center within Region XI (minimum of 1 year from the date of award)
Delivery Period: 7 working days	Delivery Period: 30 calendar days

Reference is BAC Resolution No. 729 dated November 04, 2020. For guidance and information of all concerned.


RAUL D. RAUT, ENP
(PGDH-HRMO)
BAC Chairperson