



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

BIDDING DOCUMENTS

Supply/ Delivery of Medical Equipment

PUBLIC BIDDING IB NO. G- 136-22



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Section I.
Invitation to Bid



Republic of the Philippines
PROVINCE OF DAVAO DEL SUR

Rm. 4 Executive Building, Barangay Matti, Digos City
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579 / Website: www.davaodelsur.gov.ph
Email Add.: bac.davaodelsur2@gmail.com

PGO - BIDS AND AWARDS COMMITTEE (BAC)

INVITATION TO BID FOR Supply/Delivery of Medical Equipment

1. The Provincial Government of Davao del Sur, through the 20% Dev't. Fund intends to apply the sum of ₱ 2,686,000.00 being the ABC to payments under the contract for IB No. G-136-22 (PHO). Bids received in excess of the ABC shall be automatically rejected at bid opening.
2. The Provincial Government of Davao del Sur now invites bids for the above Procurement Project. Delivery of the Goods is required by: 7 calendar days. Bidders should have completed, for the last 3 years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).
3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary "*pass/fail*" criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.
 - Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA No. 5183.
4. Prospective Bidders may obtain further information from Provincial Government of Davao del Sur and inspect the Bidding Documents at the address given below during 8:00 A.M. – 4:00 P.M., Monday to Friday.

MS. NORJANNA M. CAMAGUIN, MPA
Head of BAC Secretariat
Office of the BAC Secretariat
Room 4, Executive Building, Capitol Matti, Digos City

5. A complete set of Bidding Documents may be acquired by interested Bidders on March 24, 2022 – April 12, 2022 at the Office of the BAC Secretariat Room 4, Executive Building, Capitol Matti, Digos City upon accomplishing a bidder's assessment slip pursuant to the latest Guidelines issued by the GPPB, in the amount of ₱ 5,000.00. The Provincial Government of Davao del Sur shall allow the bidder to present its proof of payment for the fees upon payment of a non-refundable fee for the Bidding Documents in the amount stated above to the Provincial Treasurer's Office. An authorization from the proprietor of the company shall be presented in case the latter opted to send representative/s in his/her behalf. Only bidders who purchased the Bidding Documents with corresponding official receipt of the bidding documents will be allowed to submit bids. Eligibility/Technical and Financial documents must be book bound and properly tab at right side. The bidder will prepare (3) sets of book bound bidding documents with corresponding proper markings (Original Copy, Copy 1 and Copy 2), respectively. Any document submitted, not book bound shall be rejected outright. For the Financial Proposals not exceeding ten (10) pages may be book bound or securely stapled.



Republic of the Philippines
PROVINCE OF DAVAO DEL SUR

Rm. 4 Executive Building, Barangay Matti, Digos City
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579 / Website: www.davaodelsur.gov.ph
Email Add.: bac.davaodelsur2@gmail.com

PGO - BIDS AND AWARDS COMMITTEE (BAC)

6. The **Provincial Government of Davao del Sur** will hold a Pre-Bid Conference on **March 31, 2022 at 10:00 A.M. via Google Meet Application through this link: meet.google.com/ajm-bjbx-yjd**, which shall be open to prospective bidders.
7. Bids must be duly received by the Office of the BAC Secretariat, Room 4, Executive Building, Capitol Matti, Digos City through manual submission on or before **April 12, 2022 at 9:00 A.M.** Late bids shall not be accepted.
8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in **ITB Clause 14**.
9. Bid opening shall be on **April 12, 2022 at 10:00 A.M. via Google Meet Application through this link: meet.google.com/ajm-bxji-yjd**. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
10. The **Provincial Government of Davao del Sur** reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.
11. For further information, please refer to:


The BAC Chairperson
Provincial Government of Davao del Sur
Email: bac.davaodelsur2@gmail.com

MS. NORJANNA M. CAMAGUIN, MPA
Head, BAC Secretariat
Office of the BAC Secretariat
Rm. 4 Executive Building, Barangay Matti, Digos City
Email Address: bac.davaodelsur2@gmail.com
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579

12. You may visit the following websites:

For downloading of Bidding Documents : www.davaodelsur.gov.ph/

March 23, 2022


RAUL D. RAUT, ENP
(PGDH-HRMO)
BAC Chairperson



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

***Section II.
Instructions to
Bidders***



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

1. Scope of Bid

The **Provincial Government of Davao del Sur** wishes to receive Bids for the **Supply/ Delivery of Medical Equipment** with identification number **G-136-22**.

The Procurement Project (referred to herein as “**Supply/ Delivery of Medical Equipment**”) is composed of **By Item**, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for **2022** in the amount of **P2,686,000.00**.
- 2.2. The source of funding is **20 % DEV'T FUND 2022**:

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex “I” of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. *[Select one, delete other/s]*



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

- a. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
 - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
 - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
 - iii. When the Goods sought to be procured are not available from local suppliers; or
 - iv. When there is a need to prevent situations that defeat competition or restrain trade.
 - b. Foreign ownership limited to those allowed under the rules may participate in this Project.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:

[Select one, delete the other/s]

- a. For the procurement of Non-expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- b. For the procurement of Expendable Supplies: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least twenty-five percent (25%) of the ABC.
- c. For procurement where the Procuring Entity has determined, after the conduct of market research, that imposition of either (a) or (b) will likely result to failure of bidding or monopoly that will defeat the purpose of public bidding; the Bidder should comply with the following requirements: *[Select either failure or monopoly of bidding based on market research conducted]*
 - i. Completed at least two (2) similar contracts, the aggregate amount of which should be equivalent to at least *fifty percent (50%) in the*



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

case of non-expendable supplies and services or twenty-five percent (25%) in the case of expendable supplies] of the ABC for this Project; and

- ii. The largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required above.

5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

7. Subcontracts

7.1. **Subcontracting is not allowed.**

8. Pre-Bid Conference

The Provincial Government of Davao del Sur will hold a pre-bid conference on **March 31, 2022 at 10:00 A.M. via Google Meet Application through this link: meet.google.com/ajm-bjbx-yjd** as indicated in paragraph 6 of the **IB**.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *[state relevant period as provided in paragraph 2 of the IB]* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.
- 11.5. *[Include if Framework Agreement will be used:]* Financial proposals for single or multi-year Framework Agreement shall be submitted before the deadline of submission of bids as prescribed in the **IB**. For multi-year Framework Agreement, evaluation of the financial proposal during this stage is for purposes of determining eligibility and whether or not such financial proposal is within the ABC.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable);
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in e.
 - b. For Goods offered from abroad:



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

- i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
 - ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications)**.
- 12.2. *[Include if Framework Agreement will be used:]* For Framework Agreement, the following should also apply in addition to Clause 12.1:
- a. For a single year Framework Agreement, the prices quoted by the Bidder shall be fixed during the Bidder's performance of the contract and not subject to variation or escalation on any account. Price schedules required under Clause 12.1 shall be submitted with the bidding documents.
 - b. For a multi-year Framework Agreement, the prices quoted by the Bidder during submission of eligibility documents shall be the ceiling and the price quoted during mini-competition must not exceed the initial price offer. The price quoted during call for mini-competition shall be fixed during the Bidder's performance of that Call-off and not subject to variation or escalation on any account. Price schedules required under Clause 12.1 shall be submitted with the bidding documents.

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in: **Philippine Pesos**.

14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security shall be valid until *[indicate date]*. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

14.3. *[Include if Framework Agreement will be used:]* In the case of Framework Agreement, other than the grounds for forfeiture under the 2016 revised IRR, the bid security may also be forfeited if the successful bidder fails to sign the Framework Agreement, or fails to furnish the performance security or performance securing declaration. Without prejudice on its forfeiture, bid securities shall be returned only after the posting of performance security or performance securing declaration, as the case may be, by the winning Bidder or compliant Bidders and the signing of the Framework Agreement.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

16.1. Bids must be delivered to the Office of the BAC Secretariat Room 4, Executive Building, Capitol Matti, Digos City on or before **April 12, 2022 at 9:00 A.M.** Late bids shall not be accepted. All bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 18.

17. Opening and Preliminary Examination of Bids

17.1. Bid Opening shall be on **April 12, 2022 at 10:00 A.M.** via Google Meet Application through this link: meet.google.com/ajm-bxbj-yjd. Bids will be opened in the presence of bidders' representatives who choose to attend the activity.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

18.2. *[Include if Framework Agreement will be used:]* For multi-year Framework Agreement, determination of margin of preference shall be conducted every call for Mini-Competition.



Republic of the Philippines
PROVINCE OF DAVAO DEL SUR

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated “*passed*,” using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as:

One Project having several items that shall be awarded as one contract.
- 19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

- 20.1. *[Include if Framework Agreement will be used:]* For multi-year Framework Agreement, all bidders initially determined to be eligible and financially compliant shall be subject to initial post-qualification. The BAC shall then recommend the execution of a Framework Agreement among all eligible, technically and financially compliant bidders and the Procuring Entity and shall be issued by HOPE a Notice to Execute Framework Agreement. The determination of the Lowest Calculated Bid (LCB) shall not be performed by the BAC until a Mini-Competition is conducted among the bidders who executed a Framework Agreement. When a Call for Mini-Competition is made, the BAC shall allow the bidders to submit their best financial proposals on such pre-scheduled date, time and place to determine the bidder with the LCB.
- 20.2. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, *{[Include if Framework Agreement will be used:]}* or in the case of multi-year Framework Agreement, that it is one of the eligible bidders who have submitted bids that are found to be technically and financially compliant,}the Bidder shall submit its latest



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**. *{[Include if Framework Agreement will be used:] For every mini-competition in Framework Agreement, the LCB shall likewise submit the required documents for final Post Qualification.}*



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Section III.
Bid Data Sheet



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Bid Data Sheet

ITB Clause	
5.3	<p>For this purpose, contracts similar to the Project shall be:</p> <p>a. Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid for the last 3 years, and whose value, adjusted to current prices using the Philippine Statistics Authority (PSA) consumer price indices, must be at least fifty percent (50%) of the ABC. However, in the case of Expendable Supplies, said SLCC must be at least twenty five percent (25%) of the ABC. Other conditions set forth under Section 23.4.1.3 of the 2016 Revised IRR of R.A. 9184 shall also be observed.</p>
7.1	<i>Subcontracting is not allowed.</i>
12	The price of the Goods shall be quoted DDP [<i>state place of destination</i>] or the applicable International Commercial Terms (INCOTERMS) for this Project.
14.1	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <p>a. The amount of not less than P53,720.00, if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</p> <p>b. The amount of not less than P134,300.00 if bid security is in Surety Bond.</p>
19.3	<i>Please see Purchase Request No. 0554-CB-22 dated March 21, 2022.</i>
20.2	<i>Post-Qualification Requirements</i>
21.2	<p>Each Bidder shall submit:</p> <p>1.Envelope 1 (Eligibility/Technical Documents) One (1) ORIGINAL COPY, another copy for COPY 1, and another copy for COPY 2.</p> <p>2. Envelope 2 (Financial Proposal) One (1) ORIGINAL COPY, another copy for COPY 1, and another copy for COPY 2</p> <p>Such documents shall be duly signed by the bidders or its duly authorized representative/s.</p> <p>Eligibility/Technical and Financial Proposal must be book bound and properly tab at the right side. Any documents submitted not book bound shall be rejected outright. For the Financial Proposal not exceeding ten (10) pages maybe book bound or securely stapled.</p> <p>All copies must be properly marked (ORIGINAL COPY, COPY 1 AND COPY 2)</p>



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

	<p>Envelope 1 and Envelope 2 must be properly sealed. Envelope 1 and Envelope 2 shall be placed in one mother envelope.</p>
--	--



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

*Section IV. General
Conditions of
Contract*



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

[Include the following clauses if Framework Agreement will be used:]

- 2.3. For a single-year Framework Agreement, prices charged by the Supplier for Goods delivered and/or services performed under a Call-Off shall not vary from the prices quoted by the Supplier in its bid.
- 2.4. For multi-year Framework Agreement, prices charged by the Supplier for Goods delivered and/or services performed under a Call-Off shall not vary from the prices quoted by the Supplier during conduct of Mini-Competition.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184. *{[Include if Framework Agreement will be used:]} In the case of Framework Agreement, the*



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Bidder may opt to furnish the performance security or a Performance Securing Declaration as defined under the Guidelines on the Use of Framework Agreement.}

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project *{[Include if Framework Agreement will be used:]*or Framework Agreement} specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the SCC, **Section IV (Technical Specifications)** shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

- 6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.
- 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

***Section V.
Special Conditions of
Contract***



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Special Conditions of Contract

GCC Clause	
1	<p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).</p> <p>For purposes of this Clause the Procuring Entity's Representative at the Project Site is PGSO Warehouse.</p> <p>Incidental Services –</p> <p>The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements:</p> <p><i>Select appropriate requirements and delete the rest.</i></p> <ul style="list-style-type: none">a. performance or supervision of on-site assembly and/or start-up of the supplied Goods;b. furnishing of tools required for assembly and/or maintenance of the supplied Goods;c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and <p><i>Select appropriate requirements and delete the rest.</i></p> <ul style="list-style-type: none">e. performance or supervision of on-site assembly and/or start-up of the supplied Goods;f. furnishing of tools required for assembly and/or maintenance of the supplied Goods;g. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;h. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

- i. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.
- j. *[Specify additional incidental service requirements, as needed.]*

The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

Spare Parts –

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

Select appropriate requirements and delete the rest.

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
 - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
 - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI (Schedule of Requirements)** and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of *[indicate here the time period specified. If not used indicate a time period of three times the warranty period]*.

Spare parts or components shall be supplied as promptly as possible, but in any case, within *[insert appropriate time period]* months of placing the order.



Republic of the Philippines
PROVINCE OF DAVAO DEL SUR

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

	<p>Packaging –</p> <p>The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.</p> <p>The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.</p> <p>The outer packaging must be clearly marked on at least four (4) sides as follows:</p> <p>Name of the Procuring Entity Name of the Supplier Contract Description Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications</p>
	<p>A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.</p> <p>Transportation –</p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.</p> <p>Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.</p>



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

	<p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p> <p>Intellectual Property Rights –</p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
2.2	The terms of payment shall be Credit Basis.
4	The inspections and tests that will be conducted are: <i>[Indicate the applicable inspections and tests]</i>



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Section VI. Schedule of Requirements

PGO - BIDS AND AWARDS COMMITTEE

Schedule of Requirements

The delivery schedule as per IB NO.: G-136-22 is 7 Calendar Days

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
PR# 0554-CB-22 (PHO)					
1	<p>Heavy Duty Transport Stretcher Hospital Bed <i>Full Cart Size</i> Full length: Maximum of 1930mm Full width: Maximum of 640mm</p> <p><i>Bed base height range</i> Maximum Height: 900mm Minimum Height: 500mm</p> <p><i>Back section panel: 0 degrees to maximum of 70 degrees</i> Loading bearing up to 250kg</p> <p>*Centrally controlled brake system stable and reliable *ABS side rail *Back rest section supported pneumatic spring *Casters with brake treadle *IV pole and drainage bag hangers *Stowed oxygen bottle holder</p> <p>Accessories: *IV Pole *Mattress</p>	30	unit		
2	<p>ECG Machine 3 channels 7-inch high resolution color touch screen 12-lead simultaneous acquisition and display ECG automatic measurement and interpretation function Complete digital filters, resisting baseline drift, AC and EMG interference Support USB flash disk and micro SD card to extend memory Software upgrade via USB/SD Card Built-in rechargeable Li-ion Battery</p> <p>Lead: Standard 12 leads Acquisition Mode: Simultaneous 12 leads acquisition Measurement Range: $\pm 5\text{mVpp}$ Input Circuit: Floating; Protection circuit against Defibrillator effect Input Impedance: greater than or equal to 50M Ohms Input circuit current: $<0.05\text{micro Ampere}$ Record Mode: Automatic: 3CHx4+1R, 3CHx4, 3CHx2+2CHx3, 3CHx2+2CHx3+1R, 6CHx2; Manual: 3CH, 2CH, 3CH+1R, 2CH+1R; Rhythm: Any lead selectable Filter: EMG Filter: 25Hz/30Hz/40Hz/75Hz/100Hz/150Hz DFT Filter: 0.05Hz/0.15Hz AC Filter: 50Hz/60Hz CMRR: $>100\text{dB}$ Patient leak current: $<10\text{micro Ampere (220V-240V)}$ Frequency Response: 0.05Hz~150Hz (-3dB) Sensitivity: 2.5 mm/mV, 5 mm/mV, 10 mm/mV, 20 mm/mV (error $\pm 5\%$)</p> <p>Anti-baseline Drift: Automatic Time constant: greater than or equal to 3.2s Noise level: $<15\text{microVp-p}$ Paper speed: 12.5 mm/s, 25 mm/s, 50 mm/s Recording mode: Thermal printing system Paper specification: Roll 80mm x 20M Safety Standard: IEC I/CF Sample Rate: Normal: 1000sps/channel</p>	2	unit		

PGO - BIDS AND AWARDS COMMITTEE

Schedule of Requirements

The delivery schedule as per IB NO.: G-136-22 is 7 Calendar Days

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
	Power supply: AC: 100~240V, 50/60Hz, 30VA~100VA DC: 14.8V/2200mAh, built-in lithium battery				
3	<p>Mc-intosh Fiber Optic Laryngoscope Adult</p> <p><i>*Halogen Illumination</i></p> <p><i>*Manufactured from AISI 304 Stainless-Steel</i></p> <p><i>*Halogen Illumination 2.5V</i></p> <p><i>*Bulb is inside the handle in order to simplify the sterilization of the blades.</i></p> <p><i>*Blades and handles are interchangeable with other leading manufacturer</i></p> <p><i>*With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case</i></p>	3	unit		
4	<p>Mc-intosh Fiber Optic Laryngoscope Pediatric</p> <p><i>*Halogen Illumination</i></p> <p><i>*Manufactured from AISI 304 Stainless-Steel</i></p> <p><i>*Halogen Illumination 2.5V</i></p> <p><i>*Bulb is inside the handle in order to simplify the sterilization of the blades.</i></p> <p><i>*Blades and handles are interchangeable with other leading manufacturer</i></p> <p><i>*With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case</i></p>	3	unit		
5	<p>Miller Fiber Optic Laryngoscope Neonate</p> <p><i>*Halogen Illumination</i></p> <p><i>*Manufactured from AISI 304 Stainless-Steel</i></p> <p><i>*Halogen Illumination 2.5V</i></p> <p><i>*Bulb is inside the handle in order to simplify the sterilization of the blades.</i></p> <p><i>*Blades and handles are interchangeable with other leading manufacturer</i></p> <p><i>*With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case</i></p>	3	unit		
6	<p>Procedure light, portable floor model (5-bulb LED)</p> <p>Illumination: greater than or equal to 50,000LX (at 1m distance)</p> <p>Color temperature: 4000K±500K</p> <p>Mains voltage: AC220V±10% 50Hz</p> <p>Bulb voltage: AC24V</p> <p>Bulb power Input power: 120VA</p>	2	unit		
7	<p>Tabletop Pulse Oximeter (Heavy Duty)</p> <p><i>*7-inch color LCD display</i></p> <p><i>*SpO2 and Pulse rate measurement</i></p> <p><i>*Patient information management</i></p> <p><i>*Powerful data storage capacity</i></p> <p><i>*Remote software update</i></p> <p><i>*Audible and visible alarm capability</i></p> <p>Measurement range: SpO2: 0-100%; Pulse Rate: 20-350 (bpm)</p> <p>Resolution: Saturation (%SpO2): 1%; Pulse Rate (PR): 1bpm</p> <p>Electrical: AC Power Requirements: 100-240VAC, 47-63Hz; Power Consumption: 20VA</p> <p>Batteries: Type: Lithium Battery; Operating Time: >8 hours; Charge Time: <6 hours</p> <p>Operating Temperature: 0-45 degree Celcius</p> <p>Storage Temperature: -20-60 degree Celcius</p> <p>Operating Humidity: 30-95%</p> <p>Operating Altitude: -500-5000m</p> <p>Alarms: Audible and visual alarms for high and low saturation (0-100%), pulse rate (30-250bpm), Sensor off, full storage and low battery</p> <p>Alarm Volume: 85dB</p> <p>Accessories:</p>	3	unit		

PGO - BIDS AND AWARDS COMMITTEE

Schedule of Requirements

The delivery schedule as per IB NO.: G-136-22 is 7 Calendar Days

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
	<ul style="list-style-type: none"> *Main Unit *Adult Reusable SpO2 sensor *Power Cord *Operations Manual *Various SpO2 sensors *SpO2 Extension Cable 				
8	4-drawer Multipurpose trolley - Stainless steel structure - Upper ABS platform on top with extended up rims on three sides - Drawers with automatic shutting system, 110% total opening - Bumper protection - 125mm rolling wheels, two with brakes - Equipped with optional accessories - Dimensions 62x53x100cm or any manufacturer's equivalent	1	unit		
9	Dressing Trolley/Instrument Trolley -Stainless steel structure -Stainless top and body -Mounted on 4 pcs 2" caster wheels included in dressing cart -1 small tray -1 small basin -8 small bottles -5 big bottles -2 drawers -1 cabinet	3	unit		
10	Syringe Infusion Pump Specifications: Dimension: 258 x 75 x 152 mm (WxHxD) Weight: 1.7kg including battery Screen: 3.0inch LCD touchscreen, brightness 10 level adjustable, 7 colors selectable Classification: Type CF, Classic I, IP23 Compatible Syringe Set: All syringes of 2ml, 3ml, 5ml, 10ml, 20/30ml, 50/60ml conform to the standard Special Functions Recent therapy: 20 recent therapies are recorded and can be used for rapid infusion Rate change during infusion: Infusion rate can be changed anytime during infusion, not necessary to stop infusion before change Relay Function: Continuous infusion among any pump within docking station at any mode to guarantee non-stop therapies for sensitive drugs Auto screen lock: "OFF" and "ON" 15s, 30s, 1min, 2min, 5min, 10min, 30min selectable when "ON"."OFF" by default Drug library: More than 1500 drugs available with soft and hard limit editable (DERS), up to 5000 drugs can be built in Repeat alarming: Alarm sounds again in 2 minutes, if there is still alarm after alarm is mute Event recording: Maximum 2000 events can be stored and play back Sound volume: 5 levels selectable Power supply switching: When AC/DC power supply is cut off, the infusion automatically switch to internal battery supply	1	unit		
11	Volumetric Infusion Pump Flow rate range: 1~1000ml Delivery Volume Limit: 1~9999ml Total infused volume (Volume to be infused): 1~9999ml Accuracy with approved IV set: Within ±5% Power requirements: AC110/230V, 50/60Hz or DC9.6C, DC12V/24V, 2A	1	unit		

PGO - BIDS AND AWARDS COMMITTEE

Schedule of Requirements

The delivery schedule as per IB NO.: **G-136-22 is Calendar Days**

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
	<p>Battery type: Ni-Cd rechargeable battery Battery operating time: 4 hours (at 125ml/h) Operating condition: 10~40 degree Celcius, 30~85% RH (no condensation) Storage condition: -10~45 degree Celcius, 10~95% RH (no condensation)</p> <p>Infusion principle: Linear Peristaltic Finger Alarms (Audio & Visual): Air in line, Occlusion, Flow Error (Free Flow), Standby, K.V.O. (infusion Completion), Door open, Low battery, Device malfunction Special function: Keypad lock, Retain memory, Remaining time, Alram repeat, Open system, K.V.O., Purge, Bolus, Occ. Level, History call-back, Dosage mode (option), Nurse call (Option), Profile (Option), Central system (Option) Weight Approximately: 3.5KG (without rechargeable battery 3KG) Dimension: 100 x 190 x 250 (mm) or manufacturer's equivalent</p>				
12	<p>Examining Table with Stirrups *Examination and treatment table *Hamilton Type *With drawer and compartment door *Upholstered three section top, provided with stirrup and sliding foot stool *Size: 58 in x 20 in x 34 in or manufacturer's equivalent</p>	1	unit		
13	<p>Defibrillator</p> <ul style="list-style-type: none"> · 4-in-one design including: monitoring, Manual Defib, AED and pacer · Power on within 2s, Charging to 200J within 3s, ECG recovery within 2.5s · Up to 360J escalating dose energy to maximize defibrillation success · Powerful power capacity with battery to support continuous long time monitoring and shocks during transport without external power supply <p>*Waveform: Biphasic truncaed exponential waveform; with impedance compensation *Energy Accuracy: ± 2 J or 15% of setting, whichever is greater, into 50 Ohm *Power on time: Less than 2 seconds with a new, fully charged battery *Charge Time: Less than 3 seconds to 200 Joules with a new, fully charged battery; Less than 7 seconds to 360 Joules with a new, fully charged battery *ECG recovery time: Less than 2.5 seconds *Shock delivery: via multifuction defib electrode pads or paddles *Patient impedance range: 25 to 300 Ohm (External defibrillation) *Capable of Manual Mode and AED Mode *Includes recorder, data storage, noninvasive pacing, ECG monitoring, Respiration, SpO2 Pulse Oximetry; Temperature; NIBP; CO2 (Sidestream and microstream), IBP, and CPR compression functions.</p> <p>Technical Specifications: *Physical -Dimensions: Without external paddles: 295 mm (w) x 218 mm (d) x 279 mm (h) or Manufacturer's equivalent; With external paddles: 295 mm (w) x 218 mm (d) x 323 mm (h) or Manufacturer's equivalent</p> <p>*Display -Type:TFT Color LCD -Display: 8.4 inch or manufacturer's equivalent -Resolution: 800x600 pixels -Display Waveforms: Maximum of 4 channels -Wave Viewing Time: Maximum of 16 s (ECG)</p> <p>*Power -AC Power: Line voltage: 100 to 240V \sim ($\pm 10\%$); Current: 1.8 to 0.8 A;</p>	1	unit		

PGO - BIDS AND AWARDS COMMITTEE

Schedule of Requirements

The delivery schedule as per IB NO.: **G-136-22 is 7 Calendar Days**

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
	Frequency: 50/60 Hz (± 3 Hz) -DC Power (through DC-AC Inverter): Input voltage: 12VDC; Power Consumption: 190W -Battery: Type: 15.1V, 5600mAh, rechargeable lithium ion battery pack; Capacity Indicator: 5-segment led indicator for fast battery capacity evaluation; Capacity: Monitoring mode: 12 hours, configured with 5-lead ECG, Resp, SpO2, 2-channel temp, CO2 and NIBP measurements set at an interval of 15 minutes. Defib mode: 400 times, 360J discharge at intervals of 1 minute without recording Pacing mode: 9 hours, 50 Ohm load impedance, Pacing rate: 80bpm, Pacing Output: 60mA				
	<i>Above items should have at least 1 year warranty on parts and services; after sales service 24/7 and on-call if problem arises as needed; has services center within Region XI and General Santos City; Delivery Period: 7 calendar days.</i>				



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Section VII. Technical Specifications

PGO - BIDS AND AWARDS COMMITTEE

Technical Specifications

IAEB NO.: _____ Date of Opening: April 12, 2022 Quotation: By Lot

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	PR# 0554-CB-22 (PHO)				
1	Heavy Duty Transport Stretcher Hospital Bed <i>Full Cart Size</i> Full length: Maximum of 1930mm Full width: Maximum of 640mm <i>Bed base height range</i> Maximum Height: 900mm Minimum Height: 500mm <i>Back section panel: 0 degrees to maximum of 70 degrees</i> Loading bearing up to 250kg *Centrally controlled brake system stable and reliable *ABS side rail *Back rest section supported pneumatic spring *Casters with brake treadle *IV pole and drainage bag hangers *Stowed oxygen bottle holder Accessories: *IV Pole *Mattress		30	unit	
2	ECG Machine 3 channels 7-inch high resolution color touch screen 12-lead simultaneous acquisition and display ECG automatic measurement and interpretation function Complete digital filters, resisting baseline drift, AC and EMG interference Support USB flash disk and micro SD card to extend memory Software upgrade via USB/SD Card Built-in rechargeable Li-ion Battery Lead: Standard 12 leads Acquisition Mode: Simultaneous 12 leads acquisition Measurement Range: ±5mVpp Input Circuit: Floating; Protection circuit against Defibrillator effect Input Impedance: greater than or equal to 50M Ohms Input circuit current: <0.05micro Ampere Record Mode: Automatic: 3CHx4+1R, 3CHx4, 3CHx2+2CHx3, 3CHx2+2CHx3+1R, 6CHX2; Manual: 3CH, 2CH, 3CH+1R, 2CH+1R; Rhythm: Any lead selectable Filter: EMG Filter: 25Hz/30Hz/40hz/75Hz/100Hz/150Hz DFT Filter: 0.05Hz/0.15Hz AC Filter: 50Hz/60Hz CMRR: >100dB Patient leak current: <10micro Ampere (220V-240V) Frequency Response: 0.05Hz~150Hz (-3dB) Sensitivity: 2.5 mm/mV, 5 mm/mV, 10 mm/mV, 20 mm/mV (error ±5%) Anti-baseline Drift: Automatic Time constant: greater than or equal to 3.2s Noise level: <15microVp-p Paper speed: 12.5 mm/s, 25 mm/s, 50 mm/s Recording mode: Thermal printing system		2	unit	

PGO - BIDS AND AWARDS COMMITTEE

Technical Specifications

IAEB NO.: _____ Date of Opening: April 12, 2022 Quotation: By Lot

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	Paper specification: Roll 80mm x 20M Safety Standard: IEC I/CF Sample Rate: Normal: 1000sps/channel Power supply: AC: 100~240V, 50/60Hz, 30VA~100VA DC: 14.8V/2200mAh, built-in lithium battery				
3	Mc-intosh Fiber Optic Laryngoscope Adult <i>*Halogen Illumination</i> *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case		3	unit	
4	Mc-intosh Fiber Optic Laryngoscope Pediatric <i>*Halogen Illumination</i> *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case		3	unit	
5	Miller Fiber Optic Laryngoscope Neonate <i>*Halogen Illumination</i> *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case		3	unit	
6	Procedure light, portable floor model (5-bulb LED) Illumination: greater than or equal to 50,000LX (at 1m distance) Color temperature: 4000K±500K Mains voltage: AC220V±10% 50Hz Bulb voltage: AC24V Bulb power Input power: 120VA		2	unit	
7	Tabletop Pulse Oximeter (Heavy Duty) *7-inch color LCD display *SpO2 and Pulse rate measurement *Patient information management *Powerful data storage capacity *Remote software update *Audible and visible alarm capability Measurement range: SpO2: 0-100%; Pulse Rate: 20-350 (bpm)		3	unit	

PGO - BIDS AND AWARDS COMMITTEE

Technical Specifications

IAEB NO.: _____ Date of Opening: April 12, 2022 Quotation: By Lot

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	<p>Resolution: Saturation (%SpO2): 1%; Pulse Rate (PR): 1bpm Electrical: AC Power Requirements: 100-240VAC, 47-63Hz; Power Consumption: 20VA Batteries: Type: Lithium Battery; Operating Time: >8 hours; Charge Time: <6 hours Operating Temperature: 0-45 degree Celcius Storage Temperature: -20-60 degree Celcius Operating Humidity: 30-95% Operating Altitude: -500-5000m Alarms: Audible and visual alarms for high and low saturation (0-100%), pulse rate (30-250bpm), Sensor off, full storage and low battery Alarm Volume: 85dB</p> <p>Accessories: *Main Unit *Adult Reusable SpO2 sensor *Power Cord *Operations Manual *Various SpO2 sensors *SpO2 Extension Cable</p>				
8	<p>4-drawer Multipurpose trolley - Stainless steel structure - Upper ABS platform on top with extended up rims on three sides - Drawers with automatic shutting system, 110% total opening - Bumper protection - 125mm rolling wheels, two with brakes - Equipped with optional accessories - Dimensions 62x53x100cm or any manufacturer's equivalent</p>		1	unit	
9	<p>Dressing Trolley/Instrument Trolley -Stainless steel structure -Stainless top and body -Mounted on 4 pcs 2" caster wheels included in dressing cart</p> <p>-1 small tray -1 small basin -8 small bottles -5 big bottles -2 drawers -1 cabinet</p>		3	unit	
10	<p>Syringe Infusion Pump Specifications: Dimension: 258 x 75 x 152 mm (WxHxD) Weight: 1.7kg including battery Screen: 3.0inch LCD touchscreen, brightness 10 level adjustable, 7 colors selectable Classification: Type CF, Classic I, IP23 Compatible Syringe Set: All syringes of 2ml, 3ml, 5ml, 10ml, 20/30ml, 50/60ml conform to the standard Special Functions Recent therapy: 20 recent therapies are recorded and can be used for rapid infusion</p>		1	unit	

PGO - BIDS AND AWARDS COMMITTEE

Technical Specifications

IAEB NO.: _____ Date of Opening: April 12, 2022 Quotation: By Lot

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	Rate change during infusion: Infusion rate can be changed anytime during infusion, not necessary to stop infusion before change Relay Function: Continuous infusion among any pump within docking station at any mode to guarantee non-stop therapies for sensitive drugs Auto screen lock: "OFF" and "ON" 15s, 30s, 1min, 2min, 5min, 10min, 30min selectable when "ON". "OFF" by default Drug library: More than 1500 drugs available with soft and hard limit editable (DERS), up to 5000 drugs can be built in Repeat alarming: Alarm sounds again in 2 minutes, if there is still alarm after alarm is mute Event recording: Maximum 2000 events can be stored and play back Sound volume: 5 levels selectable Power supply switching: When AC/DC power supply is cut off, the infusion automatically switch to internal battery supply				
11	Volumetric Infusion Pump Flow rate range: 1~1000ml Delivery Volume Limit: 1~9999ml Total infused volume (Volume to be infused): 1~9999ml Accuracy with approved IV set: Within ±5% Power requirements: AC110/230V, 50/60Hz or DC9.6C, DC12V/24V, 2A Battery type: Ni-Cd rechargeable battery Battery operating time: 4 hours (at 125ml/h) Operating condition: 10~40 degree Celcius, 30~85% RH (no condensation) Storage condition: -10~45 degree Celcius, 10~95% RH (no condensation) Infusion principle: Linear Peristaltic Finger Alarms (Audio & Visual): Air in line, Occlusion, Flow Error (Free Flow), Standby, K.V.O. (infusion Completion), Door open, Low battery, Device malfunction Special function: Keypad lock, Retain memory, Remaining time, Alarm repeat, Open system, K.V.O., Purge, Bolus, Occ. Level, History call-back, Dosage mode (option), Nurse call (Option), Profile (Option), Central system (Option) Weight Approximately: 3.5KG (without rechargeable battery 3KG) Dimension: 100 x 190 x 250 (mm) or manufacturer's equivalent		1	unit	
12	Examining Table with Stirrups *Examination and treatment table *Hamilton Type *With drawer and compartment door *Upholstered three section top, provided with stirrup and sliding foot stool *Size: 58 in x 20 in x 34 in or manufacturer's equivalent		1	unit	
13	Defibrillator · 4-in-one design including: monitoring, Manual Defib, AED and pacer · Power on within 2s, Charging to 200J within 3s, ECG recovery within 2.5s		1	unit	

Technical Specifications

IAEB NO.:

Date of Opening: April 12, 2022

Quotation: By Lot

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	<p>· Up to 360J escalating dose energy to maximize defibrillation success</p> <p>· Powerful power capacity with battery to support continuous long time monitoring and shocks during transport without external power supply</p> <p>*Waveform: Biphasic truncaed exponential waveform; with impedance compensation</p> <p>*Energy Accuracy: ± 2 J or 15% of setting, whichever is greater, into 50 Ohm</p> <p>*Power on time: Less than 2 seconds with a new, fully charged battery</p> <p>*Charge Time: Less than 3 seconds to 200 Joules with a new, fully charged battery; Less than 7 seconds to 360 Joules with a new, fully charged battery</p> <p>*ECG recovery time: Less than 2.5 seconds</p> <p>*Shock delivery: via multifuction defib electrode pads or paddles</p> <p>*Patient impedance range: 25 to 300 Ohm (External defibrillation)</p> <p>*Capable of Manual Mode and AED Mode</p> <p>*Includes recorder, data storage, noninvasive pacing, ECG monitoring, Respiration, SpO2 Pulse Oximetry; Temperature; NIBP; CO2 (Sidestream and microstream), IBP, and CPR compression functions.</p> <p>Technical Specifications:</p> <p>*Physical</p> <p>-Dimensions: Without external paddles: 295 mm (w) x 218 mm (d) x 279 mm (h) or Manufacturer's equivalent; With external paddles: 295 mm (w) x 218 mm (d) x 323 mm (h) or Manufacturer's equivalent</p> <p>*Display</p> <p>-Type:TFT Color LCD</p> <p>-Display: 8.4 inch or manufacturer's equivalent</p> <p>-Resolution: 800x600 pixels</p> <p>-Display Waveforms: Maximum of 4 channels</p> <p>-Wave Viewing Time: Maximum of 16 s (ECG)</p> <p>*Power</p> <p>-AC Power:</p> <p>Line voltage: 100 to 240V ~ ($\pm 10\%$);</p> <p>Current: 1.8 to 0.8 A;</p> <p>Frequency: 50/60 Hz (± 3 Hz)</p> <p>-DC Power (through DC-AC Inverter):</p> <p>Input voltage: 12VDC;</p> <p>Power Consumption: 190W</p> <p>-Battery:</p> <p>Type: 15.1V, 5600mAh, rechargeable lithium ion battery pack;</p> <p>Capacity Indicator: 5-segment led indicator for fast battery capacity evaluation;</p> <p>Capacity:</p> <p>Monitoring mode: 12 hours, configured with 5-lead ECG, Resp, SpO2, 2-channel temp, CO2 and NIBP measurements set at an interval of 15 minutes.</p> <p>Defib mode: 400 times, 360J discharge at intervals of 1 minute without recording</p>				

PGO - BIDS AND AWARDS COMMITTEE

Technical Specifications

IAEB NO.:

Date of Opening: April 12, 2022

Quotation: By Lot

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	Pacing mode: 9 hours, 50 Ohm load impedance, Pacing rate: 80bpm, Pacing Output: 60mA				
	<i>Above items should have at least 1 year warranty on parts and services; after sales service 24/7 and on-call if problem arises as needed; has services center within Region XI and General Santos City; Delivery Period: 7 calendar days.</i>				



**Republic of the Philippines
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)
Telefax: (082) 553-9579/ Website: www.davaodelsur.gov.ph/pgo-bac
Email: bac.davaodelsur2@gmail.com

OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

Section VIII.
Checklist of Technical
and Financial
Documents

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____

Project ID No. _____

Page _____ of _____

1	2	3	4	5	6	7	8	9	10	
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)	
1	Heavy Duty Transport Stretcher Hospital Bed PR# 0554-CB-22 (PHO) <i>Full Cart Size</i> Full length: Maximum of 1930mm Full width: Maximum of 640mm <i>Bed base height range</i> Maximum Height: 900mm Minimum Height: 500mm <i>Back section panel: 0 degrees to maximum of 70 degrees</i> Loading bearing up to 250kg *Centrally controlled brake system stable and reliable *ABS side rail *Back rest section supported pneumatic spring *Casters with brake treadle *IV pole and drainage bag hangers *Stowed oxygen bottle holder Accessories: *IV Pole *Mattress			30						
2	ECG Machine 3 channels 7-inch high resolution color touch screen 12-lead simultaneous acquisition and display ECG automatic measurement and interpretation function Complete digital filters, resisting baseline drift, AC and EMG interference Support USB flash disk and micro SD card to extend memory		2							

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10	
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)	
3	<p>Software upgrade via USB/SD Card Built-in rechargeable Li-Ion Battery</p> <p>Lead: Standard 12 leads Acquisition Mode: Simultaneous 12 leads acquisition Measurement Range: ±5mVpp Input Circuit: Floating; Protection circuit against Defibrillator effect Input Impedance: greater than or equal to 50M Ohms Input circuit current: <0.05micro Ampere Record Mode: Automatic: 3CHx4+1R, 3CHx4, 3CHx2+2CHx3, 3CHx2+2CHx3+1R, 6CHx2; Manual: 3CH, 2CH, 3CH+1R, 2CH+1R; Rhythm: Any lead selectable Filter: EMG Filter: 25Hz/30Hz/40Hz/75Hz/100Hz/150Hz DFT Filter: 0.05Hz/0.15Hz AC Filter: 50Hz/60Hz CMRR: >100dB Patient leak current: <10micro Ampere (220V-240V) Frequency Response: 0.05Hz~150Hz (-3dB) Sensitivity: 2.5 mm/mV, 5 mm/mV, 10 mm/mV, 20 mm/mV (error ±5%)</p> <p>Anti-baseline Drift: Automatic Time constant: greater than or equal to 3.2s Noise level: <15microVp-p Paper speed: 12.5 mm/s, 25 mm/s, 50 mm/s Recording mode: Thermal printing system Paper specification: Roll 80mm x 20M Safety Standard: IEC I/CF Sample Rate: Normal: 1000sps/channel Power supply: AC: 100~240V, 50/60Hz, 30VA~100VA DC: 14.8V/2200mAh, built-in lithium battery</p>			3						
	Mc-intosh Fiber Optic Laryngoscope Adult		3							

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
	<p><i>*Halogen Illumination</i> *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case</p>								
4	<p>Mc-intosh Fiber Optic Laryngoscope Pediatric <i>*Halogen Illumination</i> *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case</p>		3						
5	<p>Miller Fiber Optic Laryngoscope Neonate <i>*Halogen Illumination</i> *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case</p>		3						

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)
6	Procedure light, portable floor model (5-bulb LED) Illumination: greater than or equal to 50,000LX (at 1m distance) Color temperature: 4000K±500K Mains voltage: AC220V±10% 50Hz Bulb voltage: AC24V Bulb power Input power: 120VA		2						
7	Tabletop Pulse Oximeter (Heavy Duty) *7-inch color LCD display *SpO2 and Pulse rate measurement *Patient information management *Powerful data storage capacity *Remote software update *Audible and visible alarm capability Measurement range: SpO2: 0-100%; Pulse Rate: 20-350 (bpm) Resolution: Saturation (%SpO2): 1%; Pulse Rate (PR): 1bpm Electrical: AC Power Requirements: 100-240VAC, 47-63Hz; Power Consumption: 20VA Batteries: Type: Lithium Battery; Operating Time: >8 hours; Charge Time: <6 hours Operating Temperature: 0-45 degree Celcius Storage Temperature: -20-60 degree Celcius Operating Humidity: 30-95% Operating Altitude: -500-5000m Alarms: Audible and visual alarms for high and low saturation (0-100%), pulse rate (30-250bpm), Sensor off, full storage and low battery Alarm Volume: 85dB		3						

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)
	Accessories: *Main Unit *Adult Reusable SpO2 sensor *Power Cord *Operations Manual *Various SpO2 sensors *SpO2 Extension Cable								
8	4-drawer Multipurpose trolley - Stainless steel structure - Upper ABS platform on top with extended up rims on three sides - Drawers with automatic shutting system, 110% total opening - Bumper protection - 125mm rolling wheels, two with brakes - Equipped with optional accessories - Dimensions 62x53x100cm or any manufacturer's equivalent		1						
9	Dressing Trolley/Instrument Trolley -Stainless steel structure -Stainless top and body -Mounted on 4 pcs 2" caster wheels included in dressing cart -1 small tray -1 small basin -8 small bottles -5 big bottles -2 drawers -1 cabinet		3						

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10	
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)	
10	Syringe Infusion Pump Specifications: Dimension: 258 x 75 x 152 mm (WxHxD) Weight: 1.7kg including battery Screen: 3.0inch LCD touchscreen, brightness 10 level adjustable, 7 colors selectable Classification: Type CF, Classic I, IP23 Compatible Syringe Set: All syringes of 2ml, 3ml, 5ml, 10ml, 20/30ml, 50/60ml conform to the standard Special Functions Recent therapy: 20 recent therapies are recorded and can be used for rapid infusion Rate change during infusion: Infusion rate can be changed anytime during infusion, not necessary to stop infusion before change Relay Function: Continuous infusion among any pump within docking station at any mode to guarantee non-stop therapies for sensitive drugs Auto screen lock: "OFF" and "ON" 15s, 30s, 1min, 2min, 5min, 10min, 30min selectable when "ON"; "OFF" by default Drug library: More than 1500 drugs available with soft and hard limit editable (DERS), up to 5000 drugs can be built in Repeat alarming: Alarm sounds again in 2 minutes, if there is still alarm after alarm is mute Event recording: Maximum 2000 events can be stored and play back Sound volume: 5 levels selectable Power supply switching: When AC/DC power supply is cut off, the infusion automatically switch to Internal battery supply		1							
11	Volumetric Infusion Pump Flow rate range: 1~1000ml		1							

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
	<p>Delivery Volume Limit: 1~9999ml Total infused volume (Volume to be infused): 1~9999ml Accuracy with approved IV set: Within ±5% Power requirements: AC110/230V, 50/60Hz or DC9.6C, DC12V/24V, 2A Battery type: Ni-Cd rechargeable battery Battery operating time: 4 hours (at 125ml/h) Operating condition: 10~40 degree Celsius, 30~85% RH (no condensation) Storage condition: -10~45 degree Celsius, 10~95% RH (no condensation)</p> <p>Infusion principle: Linear Peristaltic Finger Alarms (Audio & Visual): Air in line, Occlusion, Flow Error (Free Flow), Standby, K.V.O. (Infusion Completion), Door open, Low battery, Device malfunction Special function: Keypad lock, Retain memory, Remaining time, Alarm repeat, Open system, K.V.O., Purge, Bolus, Occ. Level, History call-back, Dosage mode (option), Nurse call (Option), Profile (Option), Central system (Option) Weight Approximately: 3.5KG (without rechargeable battery 3KG) Dimension: 100 x 190 x 250 (mm) or manufacturer's equivalent</p>								
12	<p>Examining Table with Stirrups *Examination and treatment table *Hamilton Type *With drawer and compartment door *Upholstered three section top, provided with stirrup and sliding foot stool *Size: 58 in x 20 in x 34 in or manufacturer's equivalent</p>		1						

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10	
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)	
13	Defibrillator . 4-in-one design including: monitoring, Manual Defib, AED and pacer . Power on within 2s, Charging to 200J within 3s, ECG recovery within 2.5s . Up to 360J escalating dose energy to maximize defibrillation success . Powerful power capacity with battery to support continuous long time monitoring and shocks during transport without external power supply *Waveform: Biphasic truncated exponential waveform: with impedance compensation *Energy Accuracy: ±2 J or 15% of setting, whichever is greater, into 50 Ohm *Power on time: Less than 2 seconds with a new, fully charged battery *Charge Time: Less than 3 seconds to 200 Joules with a new, fully charged battery; Less than 7 seconds to 360 Joules with a new, fully charged battery *ECG recovery time: Less than 2.5 seconds *Shock delivery: via multifunction defib electrode pads or paddles *Patient impedance range: 25 to 300 Ohm (External defibrillation) *Capable of Manual Mode and AED Mode *Includes recorder, data storage, noninvasive pacing, ECG monitoring, Respiration, SpO2 Pulse Oximetry; Temperature, NIBP, CO2 (Sidestream and microstream), IBP, and CPR compression functions. Technical Specifications: *Physical		1							

Price Schedule for Goods Offered from Within the Philippines
[shall be submitted with the Bid if bidder is offering goods from within the Philippines]

For Goods Offered from Within the Philippines

Name of Bidder _____ Project ID No. _____ Page _____ of _____

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)
	Above items should have at least 1 year warranty on parts and services; after sales service 24/7 and on-call if problem arises as needed; has services center within Region XI and General Santos City; Delivery Period: 7 calendar days.								
	Charges: CY 2021 SUPPLEMENTAL AIP NO. 1 - 20% LGDF Purchase of Medical Equipment and Other Necessary Supplies and Equipment for DSPH Emergency Room Medical Equipment 1-07-05-110 Purpose: FOR DSPH EMERGENCY ROOM USE.								

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____



PURCHASE REQUEST
PROVINCIAL GOVERNMENT OF DAVAO DEL SUR
 Mati, Digos City

PR Id.: REVISED ER MED EQUIPMENT
 Control No.:

Department: PHO P.R. No.: **0554-01-21** Date Controlled: **MAR 21 2022** Date P.R. Prepared: **03/22/2022**

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
1	unit	Heavy Duty Transport Stretcher Hospital Bed <i>Full Cart Size</i> Full length: Maximum of 1930mm Full width: Maximum of 640mm <i>Bed base height range</i> Maximum Height: 900mm Minimum Height: 500mm <i>Back section panel: 0 degrees to maximum of 70 degrees</i> Loading bearing up to 250kg *Centrally controlled brake system stable and reliable *ABS side rail *Back rest section supported pneumatic spring *Casters with brake treadle *IV pole and drainage bag hangers *Stowed oxygen bottle holder Accessories: *IV Pole *Mattress	30	50,000.00	1,500,000.00
2	unit	ECG Machine 3 channels 7-inch high resolution color touch screen 12-lead simultaneous acquisition and display ECG automatic measurement and interpretation function Complete digital filters, resisting baseline drift, AC and EMG interference Support USB flash disk and micro SD card to extend memory Software upgrade via USB/SD Card Built-in rechargeable Li-ion Battery Lead: Standard 12 leads Acquisition Mode: Simultaneous 12 leads acquisition Measurement Range: ±5mVpp Input Circuit: Floating; Protection circuit against Defibrillator effect Input Impedance: greater than or equal to 50M Ohms Input circuit current: <0.05micro Ampere Record Mode: Automatic: 3CHx4+1R, 3CHx4, 3CHx2+2CHx3, 3CHx2+2CHx3+1R, 6CHx2; Manual: 3CH, 2CH, 3CH+1R, 2CH+1R; Rhythm: Any lead selectable Filter: EMG Filter: 25Hz/30Hz/40hz/75Hz/100Hz/150Hz DFT Filter: 0.05Hz/0.15Hz AC Filter: 50Hz/60Hz CMRR: >100dB Patient leak current: <10micro Ampere (220V-240V) Frequency Response: 0.05Hz~150Hz (-3dB) Sensitivity: 2.5 mm/mV, 5 mm/mV, 10 mm/mV, 20 mm/mV (error ±5%) Anti-baseline Drift: Automatic Time constant: greater than or equal to 3.2s	2	70,000.00	140,000.00
Total :					P 2,686,000.00

9:40
MAR 23 2022
 S.

Purpose: FOR DSPH EMERGENCY ROOM USE.

Requested by:	Cash Availability:	Approved by:
Signature:	Signature:	Signature:
Printed Name: JONNA A. MASONGSONG MD, FPSMS, MAHA,	Printed Name: FARAH GEMMA V. BIDAN, CPA	Printed Name: MARC DOUGLAS IV CHAN CAGAS
Designation: MPH <i>Comd</i> PROVINCIAL HEALTH OFFICER II	Designation: Provincial Treasurer	Designation: Governor
Date:	MAR 23 2022 #702	BY AUTHORITY OF THE GOVERNOR ATTY. HERMINIO EDNZALES PROVINCIAL ATTORNEY

PROVINCIAL BUDGET OFFICER
JESSAMIE BUAT SANCHEZ, CPA

20% DEVT FUND

[Handwritten mark]

REMAINING BALANCE	85,050
LESS THIS PROJECT	2,686,200
ALLOTMENT BALANCE	2,771,050

DATE: MAR 23 2022

CONTROL NO: 889

PROVINCIAL BUDGET OFFICE

PROVINCIAL BUDGET OFFICE
CITY OF PASADENA
100 N. GARDEN ST.
PASADENA, CA 91761



PURCHASE REQUEST
 ROVINCIAL GOVERNMENT OF DAVAO DEL SUR
 Matti, Digos City

PR Id.: REVISED ER MED EQUIPMENT
 Control No.:


Department: PHO P.R. No.: **0554-06-22** Date Controlled: **MAR 21 2022** Date P.R. Prepared: **03/22/2022**

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
3	unit	Noise level: <15microVp-p Paper speed: 12.5 mm/s, 25 mm/s, 50 mm/s Recording mode: Thermal printing system Paper specification: Roll 80mm x 20M Safety Standard: IEC I/CF Sample Rate: Normal: 1000sps/channel Power supply: AC: 100~240V, 50/60Hz, 30VA~100VA DC: 14.8V/2200mAh, built-in lithium battery Mc-intosh Fiber Optic Laryngoscope Adult *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer	3	25,000.00	75,000.00
4	unit	*With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case Mc-intosh Fiber Optic Laryngoscope Pediatric *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer	3	25,000.00	75,000.00
5	unit	*With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case Miller Fiber Optic Laryngoscope Neonate *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer	3	25,000.00	75,000.00
6	unit	*With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case Procedure light, portable floor model (5-bulb LED) Illumination: greater than or equal to 50,000LX (at 1m distance) Color temperature: 4000K±500K Mains voltage: AC220V±10% 50Hz Bulb voltage: AC24V Bulb power Input power: 120VA	2	35,000.00	70,000.00
7	unit	Tabletop Pulse Oximeter (Heavy Duty) *7-inch color LCD display *SpO2 and Pulse rate measurement	3	50,000.00	150,000.00

Total : P 2,686,000.00

Purpose: FOR DSPH EMERGENCY ROOM USE.

Requested by:	Cash Availability:	Approved by:
Signature:		
Printed Name: JONNA A. MASONGSONG MD, FPSMS, MAHA,	FARAH GEMMA V. BIDAN, CPA	MARC DOUGLAS IV CHAN CAGAS
Designation: MPH PROVINCIAL HEALTH OFFICER II	Provincial Treasurer	Governor
Date:	#762 MAR 23 2022	BY AUTHORITY OF THE GOVERNOR: ATTY. HELEN R. GONZALES PROVINCIAL ADMINISTRATOR



DESSAINE BUAT-SANCHEZ, CPA
PROVINCIAL BUDGET OFFICER



PURCHASE REQUEST
 PROVINCIAL GOVERNMENT OF DAVAO DEL SUR
 Mati, Digos City

PR Id.: REVISED ER MED EQUIPMENT
 Control No.:

Department: PHO P.R. No.: 0554-0-20 Date Controlled: MAR 21 2022 Date P.R. Prepared: 03/22/2022
 Section:

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
8	unit	*Patient information management *Powerful data storage capacity *Remote software update *Audible and visible alarm capability Measurement range: SpO2: 0-100%; Pulse Rate: 20-350 (bpm) Resolution: Saturation (%SpO2): 1%; Pulse Rate (PR): 1bpm Electrical: AC Power Requirements: 100-240VAC, 47-63Hz; Power Consumption: 20VA Batteries: Type: Lithium Battery; Operating Time: >8 hours; Charge Time: <6 hours Operating Temperature: 0-45 degree Celcius Storage Temperature: -20-60 degree Celcius Operating Humidity: 30-95% Operating Altitude: -500-5000m Alarms: Audible and visual alarms for high and low saturation (0-100%), pulse rate (30-250bpm), Sensor off, full storage and low battery Alarm Volume: 85dB Accessories: *Main Unit *Adult Reusable SpO2 sensor *Power Cord *Operations Manual *Various SpO2 sensors *SpO2 Extension Cable	1	18,000.00	18,000.00
9	unit	4-drawer Multipurpose trolley - Stainless steel structure - Upper ABS platform on top with extended up rims on three sides - Drawers with automatic shutting system, 110% total opening - Bumper protection - 125mm rolling wheels, two with brakes - Equipped with optional accessories - Dimensions 62x53x100cm or any manufacturer's equivalent	3	15,000.00	45,000.00
10	unit	Dressing Trolley/Instrument Trolley -Stainless steel structure -Stainless top and body -Mounted on 4 pcs 2" caster wheels included in dressing cart -1 small tray -1 small basin -8 small bottles -5 big bottles -2 drawers -1 cabinet	1	80,000.00	80,000.00
		Syringe Infusion Pump Specifications: Dimension: 258 x 75 x 152 mm (WxHxD)	1	80,000.00	80,000.00
Total :					P 2,686,000.00

Purpose: FOR DSPH EMERGENCY ROOM USE.

Requested by:	Cash Availability:	Approved by:
Signature:		
Printed Name: JONNA A. MASONGSONG MD, FPSMS, MAHA,	FARAH GEMMA V. BIDAN, CPA	MARC DOUGLAS IV CHAN CAGAS
Designation: MPH PROVINCIAL HEALTH OFFICER II	Provincial Treasurer	Governor
Date:	#762 MAR 23 2022	BY AUTHORITY OF THE GOVERNOR



PURCHASE REQUEST
PROVINCIAL GOVERNMENT OF DAVAO DEL SUR
 Mati, Digos City

PR Id.: REVISED ER MED EQUIPMENT
 Control No.:

Department: PHO P.R. No.: **0554-0-20** Date Controlled: **MAR 21 2022** Date P.R. Prepared: **03/22/2022**

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
11	unit	<p>Weight: 1.7kg including battery Screen: 3.0inch LCD touchscreen, brightness 10 level adjustable, 7 colors selectable Classification: Type CF, Classic I, IP23 Compatible Syringe Set: All syringes of 2ml, 3ml, 5ml, 10ml, 20/30ml, 50/60ml conform to the standard Special Functions Recent therapy: 20 recent therapies are recorded and can be used for rapid infusion Rate change during infusion: Infusion rate can be changed anytime during infusion, not necessary to stop infusion before change Relay Function: Continuous infusion among any pump within docking station at any mode to guarantee non-stop therapies for sensitive drugs Auto screen lock: "OFF" and "ON" 15s, 30s, 1min, 2min, 5min, 10min, 30min selectable when "ON"."OFF" by default Drug library: More than 1500 drugs available with soft and hard limit editable (DERS), up to 5000 drugs can be built in Repeat alarming: Alarm sounds again in 2 minutes, if there is still alarm after alarm is mute Event recording: Maximum 2000 events can be stored and play back Sound volume: 5 levels selectable Power supply switching: When AC/DC power supply is cut off, the infusion automatically switch to internal battery supply</p> <p>Volumetric Infusion Pump Flow rate range: 1~1000ml Delivery Volume Limit: 1~9999ml Total infused volume (Volume to be infused): 1~9999ml Accuracy with approved IV set: Within ±5% Power requirements: AC110/230V, 50/60Hz or DC9.6C, DC12V/24V, 2A Battery type: Ni-Cd rechargeable battery Battery operating time: 4 hours (at 125ml/h) Operating condition: 10~40 degree Celcius, 30~85% RH (no condensation) Storage condition: -10~45 degree Celcius, 10~95% RH (no condensation)</p> <p>Infusion principle: Linear Peristaltic Finger Alarms (Audio & Visual): Air in line, Occlusion, Flow Error (Free Flow), Standby, K.V.O. (infusion Completion), Door open, Low battery, Device malfunction Special function: Keypad lock, Retain memory, Remaining time, Alarm repeat, Open system, K.V.O., Purge, Bolus, Occ. Level, History call-back, Dosage mode (option), Nurse call (Option), Profile (Option), Central system (Option) Weight Approximately: 3.5KG (without rechargeable battery 3KG) Dimension: 100 x 190 x 250 (mm) or manufacturer's equivalent</p>	1	83,000.00	83,000.00
12	unit	<p>Examining Table with Stirrups *Examination and treatment table *Hamilton Type *With drawer and compartment door *Upholstered three section top, provided with stirrup and sliding foot stool *Size: 58 in x 20 in x 34 in or manufacturer's equivalent</p>	1	25,000.00	25,000.00
Total :					P 2,686,000.00

Purpose: FOR DSPH EMERGENCY ROOM USE.

Requested by:	Cash Availability:	Approved by:
Signature:		
Printed Name: JONNA A. MASONGSONG MD, FPSMS, MAHA,	FARAH GEMMA V. BIDAN, CPA	MARC DOUGLAS IV CHAN CAGAS
Designation: MPH <i>Chief</i> PROVINCIAL HEALTH OFFICER II	Provincial Treasurer	Governor
Date:	#762 MAR 23 2022	BY AUTHORITY OF THE GOVERNOR,



PURCHASE REQUEST
PROVINCIAL GOVERNMENT OF DAVAO DEL SUR
 Mati, Digos City

PR Id.: REVISED ER MED EQUIPMENT
 Control No.:

Department: PHO P.R. No.: **0554** Date Controlled: **MAR 21 2022** Date P.R. Prepared: **03/22/2022**

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
13	unit	<p>Defibrillator</p> <ul style="list-style-type: none"> 4-in-one design including: monitoring, Manual Defib, AED and pacer Power on within 2s, Charging to 200J within 3s, ECG recovery within 2.5s Up to 360J escalating dose energy to maximize defibrillation success Powerful power capacity with battery to support continuous long time monitoring and shocks during transport without external power supply <p>*Waveform: Biphasic truncated exponential waveform; with impedance compensation</p> <p>*Energy Accuracy: ± 2 J or 15% of setting, whichever is greater, into 50 Ohm</p> <p>*Power on time: Less than 2 seconds with a new, fully charged battery</p> <p>*Charge Time: Less than 3 seconds to 200 Joules with a new, fully charged battery; Less than 7 seconds to 360 Joules with a new, fully charged battery</p> <p>*ECG recovery time: Less than 2.5 seconds</p> <p>*Shock delivery: via multifunction defib electrode pads or paddles</p> <p>*Patient impedance range: 25 to 300 Ohm (External defibrillation)</p> <p>*Capable of Manual Mode and AED Mode</p> <p>*Includes recorder, data storage, noninvasive pacing, ECG monitoring, Respiration, SpO2 Pulse Oximetry; Temperature; NIBP; CO2 (Sidestream and microstream), IBP, and CPR compression functions.</p> <p>Technical Specifications:</p> <p>*Physical</p> <ul style="list-style-type: none"> Dimensions: Without external paddles: 295 mm (w) x 218 mm (d) x 279 mm (h) or Manufacturer's equivalent; With external paddles: 295 mm (w) x 218 mm (d) x 323 mm (h) or Manufacturer's equivalent <p>*Display</p> <ul style="list-style-type: none"> Type: TFT Color LCD Display: 8.4 inch or manufacturer's equivalent Resolution: 800x600 pixels Display Waveforms: Maximum of 4 channels Wave Viewing Time: Maximum of 16 s (ECG) <p>*Power</p> <ul style="list-style-type: none"> AC Power: <ul style="list-style-type: none"> Line voltage: 100 to 240V $\sim (\pm 10\%)$; Current: 1.8 to 0.8 A; Frequency: 50/60 Hz (± 3 Hz) DC Power (through DC-AC Inverter): <ul style="list-style-type: none"> Input voltage: 12VDC; Power Consumption: 190W Battery: <ul style="list-style-type: none"> Type: 15.1V, 5600mAh, rechargeable lithium ion battery pack; Capacity Indicator: 5-segment led indicator for fast battery capacity evaluation; Capacity: 	1	350,000.00	350,000.00
Total :					P 2,686,000.00

Purpose: FOR DSPH EMERGENCY ROOM USE.

Requested by:	Cash Availability:	Approved by:
Signature:	Signature:	Signature:
Printed Name: JONNA A. MASONGSONG MD, FPSMS, MAHA,	Printed Name: FARAH GEMMA V. BIDAN, CPA	Printed Name: MARC DOUGLAS IV CHAN CAGAS
Designation: MPH ^{Center} PROVINCIAL HEALTH OFFICER II	Designation: Provincial Treasurer	Designation: Governor
Date:	MAR 23 2022	BY AUTHORITY OF THE GOVERNOR

TIME: 9:40
 DATE: MAR 23 2022

 ATTY. HECTOR R. GONZALES
 PROVINCIAL ADMINISTRATOR



PURCHASE REQUEST
 PROVINCIAL GOVERNMENT OF DAVAO DEL SUR
 Mati, Digos City

PR Id.: REVISED ER MED EQUIPMENT
 Control No.:

Department: PHO P.R. No.: **0554-*bn*** Date Controlled: **MAR 21 2022** Date P.R. Prepared: **03/22/2022**
 Section:

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
		<p>Monitoring mode: 12 hours, configured with 5-lead ECG, Resp, SpO2, 2-channel temp, CO2 and NIBP measurements set at an interval of 15 minutes.</p> <p>Defib mode: 400 times, 360J discharge at intervals of 1 minute without recording</p> <p>Pacing mode: 9 hours, 50 Ohm load impedance, Pacing rate: 80bpm, Pacing Output: 60mA</p> <p>Above items should have at least 1 year warranty on parts and services; after sales service 24/7 and on-call if problem arises as needed; has services center within Region XI and General Santos City; Delivery Period: 7 calendar days.</p> <p align="center">Charges: CY 2021 SUPPLEMENTAL AIP NO. 1 - 20% LGDF Purchase of Medical Equipment and Other Necessary Supplies and Equipment for DSPH Emergency Room Medical Equipment 1-07-05-110</p> <p>BIDS & AWARDS COMMITTEE (BAC) TECHNICAL WORKING GROUP</p> <p>TECHNICAL SPECS CONTROLLED BY: <i>[Signature]</i> DATE: 2022-03-22</p> <p>BIDS & AWARDS COMMITTEE (BAC) BAC SECRETARIAT</p>			
Total :					P 2,686,000.00

Purpose: FOR DSPH EMERGENCY ROOM USE.

Requested by:	Cash Availability:	Approved by:
Signature: <i>[Signature]</i>	<i>[Signature]</i>	<i>[Signature]</i>
Printed Name: JONNA A. MASONGSONG MD, FPSMS, MAHA,	FARAH GEMMA V. BIDAN, CPA	MARC DOUGLAS IV CHAN CAGAS
Designation: MPH <i>Copy</i> PROVINCIAL HEALTH OFFICER II	Provincial Treasurer	Governor
Date:	#762 MAR 23 2022	BY AUTHORITY OF THE GOVERNOR.

RECEIVED
 TIME: 9:00
 DATE: MAR 23 2022
 SIGNATURE: *[Signature]*

ATTY. HELENERA GONZALES
 PROVINCIAL ADMINISTRATOR