



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

# **BIDDING DOCUMENTS**

## *Supply/ Delivery of Medical Equipment*

**PUBLIC BIDDING IB NO. G- 245-22**

**MAY 11, 2022** *qr*

*K*



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

***Section I.***  
***Invitation to Bid***



## PGO - BIDS AND AWARDS COMMITTEE (BAC)

### INVITATION TO BID FOR Supply/Delivery of Medical Equipment

1. The Provincial Government of Davao del Sur, through the General Fund intends to apply the sum of ₱ 4,606,000.00 being the ABC to payments under the contract for IB No. G-245-22 (PHO). Bids received in excess of the ABC shall be automatically rejected at bid opening.
2. The Provincial Government of Davao del Sur now invites bids for the above Procurement Project. Delivery of the Goods is required by 30 calendar days. Bidders should have completed, for the last 3 years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).
3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary "*pass/fail*" criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.
  - Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA No. 5183.
4. Prospective Bidders may obtain further information from Provincial Government of Davao del Sur and inspect the Bidding Documents at the address given below during 8:00 A.M. – 4:00 P.M., Monday to Friday.

**MS. NORJANNA M. CAMAGUIN, MPA**  
**Head of BAC Secretariat**  
**Office of the BAC Secretariat**  
**Room 4, Executive Building, Barangay Matti, Digos City**

5. A complete set of Bidding Documents may be acquired by interested Bidders on May 12, 2022 – May 31, 2022 at the Office of the BAC Secretariat Room 4, Executive Building, Barangay Matti, Digos City upon accomplishing a bidder's assessment slip pursuant to the latest Guidelines issued by the GPPB, in the amount of ₱ 5,000.00. The Provincial Government of Davao del Sur shall allow the bidder to present its proof of payment for the fees upon payment of a non-refundable fee for the Bidding Documents in the amount stated above to the Provincial Treasurer's Office. An authorization from the proprietor of the company shall be presented in case the latter opted to send representative/s in his/her behalf. Only bidders who purchased the Bidding Documents with corresponding official receipt of the bidding documents will be allowed to submit bids. Eligibility/Technical and Financial documents must be book bound and properly tab at right side. The bidder will prepare (3) sets of book bound bidding documents with corresponding proper markings (Original Copy, Copy 1 and Copy 2), respectively. Any document submitted, not book bound shall be rejected outright. For the Financial Proposals not exceeding ten (10) pages may be book bound or securely stapled.



## PGO - BIDS AND AWARDS COMMITTEE (BAC)

6. The **Provincial Government of Davao del Sur** will hold a Pre-Bid Conference on **May 19, 2022** at 10:00 A.M. via **Google Meet Application through this link: [meet.google.com/ajm-bjbx-yjd](https://meet.google.com/ajm-bjbx-yjd)**, which shall be open to prospective bidders.
7. Bids must be duly received by the Office of the BAC Secretariat, Room 4, Executive Building, Barangay Matti, Digos City through manual submission on or before **May 31, 2022 at 9:00 A.M.** Late bids shall not be accepted.
8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 14.
9. Bid opening shall be on **May 31, 2022 at 10:00 A.M. via Google Meet Application through this link: [meet.google.com/ajm-bjbx-yjd](https://meet.google.com/ajm-bjbx-yjd)**. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.
10. The **Provincial Government of Davao del Sur** reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.
11. For further information, please refer to:


**The BAC Chairperson**  
**Provincial Government of Davao del Sur**  
**Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)**

**MS. NORJANNA M. CAMAGUIN, MPA**  
**Head, BAC Secretariat**  
**Office of the BAC Secretariat**  
**Rm. 4 Executive Building, Barangay Matti, Digos City**  
**Email Address: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)**  
**Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)**  
**Telefax: (082) 553-9579**

12. You may visit the following websites:

**For downloading of Bidding Documents : [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)**

**May 11, 2022**

  
**RAUL D. RAUT, ENP**  
**(PGDH-HRMO)**  
**BAC Chairperson**



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

***Section II.***  
***Instructions to***  
***Bidders***



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## **OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

### **1. Scope of Bid**

The **Provincial Government of Davao del Sur** wishes to receive Bids for the **Supply/Delivery of Medical Equipment** with identification number **G-245-22**.

The Procurement Project (referred to herein as "**Supply/Delivery of Medical Equipment**") is composed of **By Item**, the details of which are described in Section VII (Technical Specifications).

### **2. Funding Information**

2.1. The GOP through the source of funding as indicated below for **2022** in the amount of **P4,606,000.00**.

2.2. The source of funding is **General Fund 2022**:

### **3. Bidding Requirements**

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

### **4. Corrupt, Fraudulent, Collusive, and Coercive Practices**

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

### **5. Eligible Bidders**

5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

5.2. *[Select one, delete other/s]*

- a. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
  - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
  - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
  - iii. When the Goods sought to be procured are not available from local suppliers; or
  - iv. When there is a need to prevent situations that defeat competition or restrain trade.
- b. Foreign ownership limited to those allowed under the rules may participate in this Project.

5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:

*[Select one, delete the other/s]*

- a. For the procurement of Non-expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- b. For the procurement of Expendable Supplies: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least twenty-five percent (25%) of the ABC.
- c. For procurement where the Procuring Entity has determined, after the conduct of market research, that imposition of either (a) or (b) will likely result to failure of bidding or monopoly that will defeat the purpose of public bidding: the Bidder should comply with the following requirements: *[Select either failure or monopoly of bidding based on market research conducted]*



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

- i. Completed at least two (2) similar contracts, the aggregate amount of which should be equivalent to at least *fifty percent (50%) in the case of non-expendable supplies and services or twenty-five percent (25%) in the case of expendable supplies*] of the ABC for this Project; and
- ii. The largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required above.

5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

### 6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

### 7. Subcontracts

7.1. **Subcontracting is not allowed.**

### 8. Pre-Bid Conference

The Provincial Government of Davao del Sur will hold a pre-bid conference on **May 19, 2022 at 10:00 A.M. via Google Meet Application through this link: [meet.google.com/ajm-bjbx-yjd](https://meet.google.com/ajm-bjbx-yjd)** as indicated in paragraph 6 of the **IB**.

### 9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

### 10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *[state relevant period as provided in paragraph 2 of the IB]* prior to the deadline for the submission and receipt of bids.





**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

### 11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.
- 11.5. *[Include if Framework Agreement will be used:]* Financial proposals for single or multi-year Framework Agreement shall be submitted before the deadline of submission of bids as prescribed in the **IB**. For multi-year Framework Agreement, evaluation of the financial proposal during this stage is for purposes of determining eligibility and whether or not such financial proposal is within the ABC.

### 12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
- a. For Goods offered from within the Procuring Entity's country:
    - i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable);
    - ii. The cost of all customs duties and sales and other taxes already paid or payable;



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## **OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

- iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
  - iv. The price of other (incidental) services, if any, listed in e.
- b. For Goods offered from abroad:
- i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
  - ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications)**.
- 12.2. *[Include if Framework Agreement will be used:]* For Framework Agreement, the following should also apply in addition to Clause 12.1:
- a. For a single year Framework Agreement, the prices quoted by the Bidder shall be fixed during the Bidder's performance of the contract and not subject to variation or escalation on any account. Price schedules required under Clause 12.1 shall be submitted with the bidding documents.
  - b. For a multi-year Framework Agreement, the prices quoted by the Bidder during submission of eligibility documents shall be the ceiling and the price quoted during mini-competition must not exceed the initial price offer. The price quoted during call for mini-competition shall be fixed during the Bidder's performance of that Call-off and not subject to variation or escalation on any account. Price schedules required under Clause 12.1 shall be submitted with the bidding documents.

### **13. Bid and Payment Currencies**

13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.

13.2. Payment of the contract price shall be made in: **Philippine Pesos**.

### **14. Bid Security**



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

- 14.1. The Bidder shall submit a Bid Securing Declaration<sup>1</sup> or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security shall be valid until *[indicate date]*. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.
- 14.3. *[Include if Framework Agreement will be used:]* In the case of Framework Agreement, other than the grounds for forfeiture under the 2016 revised IRR, the bid security may also be forfeited if the successful bidder fails to sign the Framework Agreement, or fails to furnish the performance security or performance securing declaration. Without prejudice on its forfeiture, bid securities shall be returned only after the posting of performance security or performance securing declaration, as the case may be, by the winning Bidder or compliant Bidders and the signing of the Framework Agreement.

### 15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

### 16. Deadline for Submission of Bids

- 16.1. Bids must be delivered to the Office of the BAC Secretariat Room 4, Executive Building, Capitol Matti, Digos City on or before **May 31, 2022 at 9:00 A.M.** Late bids shall not be accepted. All bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 18.

### 17. Opening and Preliminary Examination of Bids

- 17.1. Bid Opening shall be on **May 31, 2022 at 10:00 A.M.** via Google Meet Application through this link: [meet.google.com/ajm-bxbj-yjd](https://meet.google.com/ajm-bxbj-yjd). Bids will be opened in the presence of bidders' representatives who choose to attend the activity.

<sup>1</sup> In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.



**Republic of the Philippines**  
**PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

- 17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

### 18. Domestic Preference

- 18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.
- 18.2. *[Include if Framework Agreement will be used:]* For multi-year Framework Agreement, determination of margin of preference shall be conducted every call for Mini-Competition.

### 19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated “*passed*,” using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as:  
  
One Project having several items that shall be awarded as one contract.
- 19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

### 20. Post-Qualification

- 20.1. *[Include if Framework Agreement will be used:]* For multi-year Framework Agreement, all bidders initially determined to be eligible and financially compliant



**Republic of the Philippines**  
**PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

shall be subject to initial post-qualification. The BAC shall then recommend the execution of a Framework Agreement among all eligible, technically and financially compliant bidders and the Procuring Entity and shall be issued by HOPE a Notice to Execute Framework Agreement. The determination of the Lowest Calculated Bid (LCB) shall not be performed by the BAC until a Mini-Competition is conducted among the bidders who executed a Framework Agreement. When a Call for Mini-Competition is made, the BAC shall allow the bidders to submit their best financial proposals on such pre-scheduled date, time and place to determine the bidder with the LCB.

- 20.2. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, *{[Include if Framework Agreement will be used:]}* or in the case of multi-year Framework Agreement, that it is one of the eligible bidders who have submitted bids that are found to be technically and financially compliant, *}* the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**. *{[Include if Framework Agreement will be used:]}* For every mini-competition in Framework Agreement, the LCB shall likewise submit the required documents for final Post Qualification. *}*



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

# *Section III. Bid Data Sheet*



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

***Bid Data Sheet***

ITB Clause	
5.3	<p>For this purpose, contracts similar to the Project shall be:</p> <p>a. Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid for the <b>last 3 years</b>, and whose value, adjusted to current prices using the Philippine Statistics Authority (PSA) consumer price indices, must be at least fifty percent (50%) of the ABC. However, in the case of Expendable Supplies, said SLCC must be at least twenty five percent (25%) of the ABC. Other conditions set forth under Section 23.4.1.3 of the 2016 Revised IRR of R.A. 9184 shall also be observed.</p>
7.1	<b><i>Subcontracting is not allowed.</i></b>
12	The price of the Goods shall be quoted DDP [ <i>state place of destination</i> ] or the applicable International Commercial Terms (INCOTERMS) for this Project.
14.1	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <p>a. The amount of not less than <b>P 92,120.00</b> , if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</p> <p>b. The amount of not less than <b>P230,300.00</b> if bid security is in Surety Bond.</p>
19.3	<i>Please see Purchase Request No. <b>0799-CB-22 dated April 29, 2022</b></i>
20.2	<i>Post-Qualification Requirements</i>
21.2	<p><b>Each Bidder shall submit:</b></p> <p><b>1. Envelope 1 (Eligibility/Technical Documents)</b> One (1) <b>ORIGINAL COPY</b>, another copy for <b>COPY 1</b>, and another copy for <b>COPY 2</b>.</p> <p><b>2. Envelope 2 (Financial Proposal)</b> One (1) <b>ORIGINAL COPY</b>, another copy for <b>COPY 1</b>, and another copy for <b>COPY 2</b></p> <p>Such documents shall be duly signed by the bidders or its duly authorized representative/s.</p> <p>Eligibility/Technical and Financial Proposal <b>must be book bound</b> and properly tab at the right side. <b>Any documents submitted not book bound shall be rejected outright.</b> For the Financial Proposal not exceeding ten (10) pages maybe book bound or securely stapled.</p> <p>All copies must be <b>properly marked (ORIGINAL COPY, COPY 1 AND COPY</b></p>

2



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

	<p><b>2)</b> <b>Envelope 1 and Envelope 2 must be properly sealed. Envelope 1 and Envelope 2 shall be placed in one mother envelope.</b></p>
--	--

2



K-640

7-14



# PURCHASE REQUEST

PROVINCIAL GOVERNMENT OF DAVAO DEL SUR  
Matti, Digos City

PR Id.: DSPH MEDICAL EQUIPMENT  
Control No.:


Department: PHO P.R. No.: **0799-CP-N** Date Controlled: **APR 29 2022** Date P.R. Prepared: **04/08/2022**

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
1	unit	<b>Clinical Height and Weighing Scale</b>  *Height rod in inches and centimeters *Platform cover is easily removed for cleaning *Die-cast beam can be easily read from either side *Max weight: 160 kg *Scale division (D): 100 g *Height range to be measured: 75-200cm *Minimum value of height per division: 0.5cm *Size of packaging box: 1500 x 510 x 155cm *Load platform area (LxW): 375 x 275 mm *Measurement(LxWxH): 530x275x1485mm *Net weight: 15kg *Gross weight: 17kg	4	17,500.00	70,000.00
2	unit	<b>CT Single Syringe Injection System, Pedestal Type with Console</b>  <i>Pedestal</i>  Overview Features: *Touch Screen/Color LCD *Simple Operation *Friendly design *Easy to understand result and easy protocol entry *Improves throughput minimizes error *Easy to learn and less training  Console: *Flow rate: 0.1-10mL/sec *Volume: 1-200mL *Pressure Limit: 10-300 PSI *Rise Time: OFF: 0.5 sec Fixed; ON: 2 sec Fixed *Real Time Pressure Graph: YES *Protocol Memory: 160 Protocols for five (5) users *Scan Time: 0-300 sec *Remote Injection Start/Abort/Hold: YES *Synchronization Features: YES  Powerhead: *Air Check Button: YES *Injection Start/Abort/Hold: YES *Syringe Adapter: 100/200 mL *Contrast Warmer: YES  <i>At least 1 year warranty on parts and services; Installation, training and decommissioning included; After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI of General Santos City; Delivery Period; 30 calendar days.</i>	1	1,880,000.00	1,880,000.00
3	unit	<b>Hospital Emergency Cart</b> <i>HLW (105 cm x 25 cm x 36 cm) (105 cm x 25 cm x 36 cm)</i> Made of Stainless Steel 3 Drawers and 2 doors	4	19,000.00	76,000.00
<b>Total :</b>					<b>P 4,606,000.00</b>

**G-245-22**  
Preproc: 05-10-22  
Prebid: 05-19-22  
Est: 05-31-22

**Purpose:** FOR DSPH IMAGING DEPARTMENT, PHARMACY DEPARTMENT, LABORATORY DEPARTMENT AND NURSING SERVICES USE.

<b>Requested by:</b> Signature: <i>[Signature]</i> Printed Name: <b>JONNA A. MASONGSONG MD, FPSMS, MAHA,</b> Designation: <b>MPH Provincial Health Officer II</b>	<b>Cash Availability:</b> <i>[Signature]</i> <b>FARAH GEMMA V. BIDAN, CPA</b> Provincial Treasurer	<b>Approved by:</b> <i>[Signature]</i> <b>MARC DOUGLAS IV CHAN CAGAS</b> GOVERNOR <b>ATTY. HERBER R. GONZALEZ</b> PROVINCIAL ADMINISTRATOR
Date: <b>APR 28 2022</b>		






ESSAMIE BUA - SANCHEZ, CPA  
PROVINCIAL BUDGET OFFICER

GENERAL FUND

2

REMAINING BALANCE:	4,894,000
LESS THIS REQUEST:	4,000,000
ALLOTMENT:	9,500,000
DATE:	APR 27 2022
FORM NO.:	1198
Provincial Budget Office	



2



**PURCHASE REQUEST**  
**PROVINCIAL GOVERNMENT OF DAVAO DEL SUR**  
 Mati, Digos City

PR Id.: DSPH MEDICAL EQUIPMENT  
 Control No.:

Department: PHO P.R No.: **0799-0-0** Date Controlled: **APR 29 2022** Date P.R. Prepared: **04/08/2022**

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
4	unit	Stainless Top IV hook with Bumper <b>Hot Plate with Magnetic Stirrer 220V</b> *5" stainless steel with ceramic coated plate *Heating temperature range 50°C - 320°C *LED speed display *Brushless DC motor *Motor Rating Output: at least 20W *Voltage 200-240/650W *Speed range 50-1500 rpm  <b>At least 1 year warranty on parts and services</b>	1	40,000.00	40,000.00
5	unit	<b>Mc-intosh Fiber Optic Laryngoscope Adult</b> *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case	5	25,000.00	125,000.00
6	unit	<b>Mc-intosh Fiber Optic Laryngoscope Pediatric</b> *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer	1	25,000.00	25,000.00
7	unit	<b>Uninterrupted Power Supply Tender for CT Scan</b>  *High frequency online type UPS used in medical equipment industry *Compatible with existing 16 Slice CT Scan machine *Must be three-phase with neutral plus ground  *Capacity: at least 80KVA *4 hours back up time and 2 hours scanning  *Nominal Voltage Input: at least 380VAC phase to phase and 220VAC phase to neutral or ground *Voltage Range: Full load: at least 305VAC phase to phase and 176VAC phase to neutral or ground; De-rating: at least 208 VAC 70 percent phase to phase and 120VAC phase to neutral or ground *Input Voltage Comeback: Low loss volatge +10V, high loss -10V *Frequency Range: 40 to 70 Hertz	1	1,900,000.00	1,900,000.00
<b>Total :</b>					<b>P 4,606,000.00</b>


**Purpose:** FOR DSPH IMAGING DEPARTMENT, PHARMACY DEPARTMENT, LABORATORY DEPARTMENT AND NURSING SERVICES USE.

Requested by: \_\_\_\_\_ Cash Availability: \_\_\_\_\_ Approved by: \_\_\_\_\_

Signature: \_\_\_\_\_  
 Printed Name: **JONNA A. MASONGSONG MD, FPSMS, MAHA,** **FARAH GEMMA V. BIDAN, CPA** **MARC DOUGLAS IV CHAN CAGAS**

Designation: **MPH** **Provincial Treasurer** **GOVERNOR OF THE GOVERNOR**  
**PROVINCIAL HEALTH OFFICER II**

Date: **APR 28 2022** **ATTY. HERBERT R. GUNZALE**  
**#1040** **PROVINCIAL ADMINISTRATOR**







**PURCHASE REQUEST**  
**PROVINCIAL GOVERNMENT OF DAVAO DEL SUR**  
 Mati, Digos City

PR Id.: DSPH MEDICAL EQUIPMENT  
 Control No.:

Department: PHO P.R No.: **0799-UN** Date Controlled: **APR 29 2022** Date P.R. Prepared: **04/08/2022**

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
		<p>*Input Synchronization Range: +/-1Hz,+/-2Hz,+/-4Hz and user programmable depends on the compatibility needs</p> <p>*Power Factor: 1.0</p> <p>*THDi at 100 percent load: Must be less than 3 percent</p> <p>*Nominal bypass voltage: at least 380VAC phase to phase and 220 phase to neutral or ground</p> <p>*Nominal bypass voltage range: +15% upper limit and -20% lower limit</p> <p>*Nominal output frequency: 60 Hz</p> <p>*Nominal output frequency range: +/-1Hz,+/-2Hz,+/-4Hz</p> <p>*Bypass synchronization range: +/-1Hz,+/-2Hz,+/-4Hz</p> <p>*Output voltages: at least 380VAC phase to phase and 230VAC phase to neutral or ground</p> <p>*AC regulation range: Less than 1% at typical balance load and less than 2% for unbalance load</p> <p>*Output frequency range: 50-60Hz</p> <p>*Frequency range on battery mode: 60Hz +/-1</p> <p>*Overload AC mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload</p> <p>*Overload battery mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload</p> <p>*Current crest ratio: 3:1 max</p> <p>*Harmonic distortion: 100% linear load should be less than 2% and 100% non linear load should be less than 4%</p> <p>*Transfer time AC Line to battery: 0ms</p> <p>*Battery Mode backup time: Must not be less than 1hour CT scan scanning mode and 4 hours on CT scan standby mode</p> <p>*Battery type: VRLA *Battery number: at least 40 pieces with rated specs of 12VDC and 150AH</p> <p>*Battery Cabinet: Must be customizable depends on the allocated space for electrical room</p> <p>*Maximum charging current: 24A</p> <p>*Floating voltage: 2.29V/cell</p> <p>*Boost charging: 2.35V/cell</p> <p>*Charger voltage regulation: Less than 1% *Ripple voltage: Less than 1%</p> <p>*Ripple current: Less than 5</p> <p>*UPS controller cabinet dimension: Should not be more than 950cmX570cmX1020cm</p> <p>*UPS controller cabinet weight: Should not be more than 195kg</p> <p>*Environmental operating temperature: 0-40 degrees Celcius</p> <p>*Environmental operating humidity: Should be less than 95% and non-condensing</p> <p>*Operating altitude: Should be less than 1000meters</p> <p>*Acoustic noise level: Should be less than 70dB at 1 meter distance</p> <p>*UPS smart system management: RS232 or USB and should support Windows, Linux and IOS operating system</p> <p>*Power management: Must have and optional SNMP module</p>			

*Handwritten notes:*  
 - Voltage Direct Current

Total : P 4,606,000.00

Purpose: FOR DSPH IMAGING DEPARTMENT, PHARMACY DEPARTMENT, LABORATORY DEPARTMENT AND NURSING SERVICES USE.

Requested by:	Cash Availability:	Approved by:
Signature:		
Printed Name: <b>JONNA A. MASONGSONG MD, FPSMS, MAHA,</b>	<b>FARAH GEMMA V. BIDAN, CPA</b>	<b>MARC DOUGLAS IV CHAN CAGAS</b>
Designation: <b>MPH <i>Canyo</i></b> <b>PROVINCIAL HEALTH OFFICER II</b>	<b>Provincial Treasurer</b>	<b>Governor</b> <b>ATTY. HERBERT R. GONZALE</b> <b>PROVINCIAL ADMINISTRATOR</b>
Date: <b>20 2022</b> <b>#1040</b>		





**PURCHASE REQUEST**  
 PROVINCIAL GOVERNMENT OF DAVAO DEL SUR  
 Mati, Digos City

PR Id.: DSPH MEDICAL EQUIPMENT  
 Control No.:

Department: PHO P.R. No.: **0799-02N** Date Controlled: **APR 29 2022** Date P.R. Prepared: **04/08/2022**  
 Section:

Item No.	Unit	Item and Description	Qty	Unit Cost	Total Cost
8	unit	<p><b>3 years warranty (1 year parts and labor + 2 years unlimited support on equipment) 1 year warranty on batteries;</b>            Installation, training and decommissioning included;            After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI <del>or</del> General Santos City; Delivery Period: 30 calendar days.</p> <p><b>Vaccine Storage Refrigerator</b>            Specs:            *Type: Upright, Double Glass Door            *Capacity: At least 940 Liters            *Temperature Range: 2 degrees Celcius to 8 degrees Celcius            *Forced air cooling, auto defrost &amp; microprocessor control            *Bright LED display            *Dimensions (WxDxH): 1130 x 755 x 1980 mm or manufacturer's standard dimensions equivalent            *Power: 208-230/60            *Good for at least 9, 450 vials            *With 1 unit of AVR 3KVA</p> <p><b>1 YEAR WARRANTY; PREVENTIVE MAINTENANCE (SEMI-ANNUAL); After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI <del>or</del> General Santos City; Delivery Period: 30 calendar days.</b></p> <p align="center">Charges:            GENERAL FUND 2022 CAPITAL OUTLAY - MEDICAL EQUIPMENT 1-07-05-110</p> <p><b>BIDS &amp; AWARDS COMMITTEE (BAC)            TECHNICAL WORKING GROUP</b>            BY: <i>[Signature]</i>            DATE: <b>2022-04-28</b></p> <p><b>BIDS &amp; AWARDS COMMITTEE (BAC)            AG SECRETARIA</b>            COVERED UNDER APP 20            CONTROLLED BY: <i>[Signature]</i>            DATE: <b>4/29/22</b>  <b>799 CB</b></p>	1	490,000.00	490,000.00
<b>Total :</b>					<b>P 4,606,000.00</b>


**Purpose:** FOR DSPH IMAGING DEPARTMENT, PHARMACY DEPARTMENT, LABORATORY DEPARTMENT AND NURSING SERVICES USE.

**Requested by:** \_\_\_\_\_ **Cash Availability:** \_\_\_\_\_ **Approved by:** \_\_\_\_\_

Signature: *[Signature]* *[Signature]* *[Signature]*  
 Printed Name: **JONNA A. MASONGSONG MD, FPSMS, MAHA,** **FARAH GENMA V. BIDAN, CPA** **MARC DOUGLAS IV CHAN CAGAS**

Designation: **MPH *[Signature]*** **Provincial Treasurer** **Governor**  
**PROVINCIAL HEALTH OFFICER II**

Date: \_\_\_\_\_ **APR 28 2022** **RECEIVED**  
*#1640* **PROVINCIAL TREASURER'S OFFICE** **GOVERNOR**  
**MATTY. HERBERT R. GONZALE**  
**PROVINCIAL ADMINISTRATOR**





**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

# *Section IV. General Conditions of Contract*



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## **OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

### **1. Scope of Contract**

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

### **2. Advance Payment and Terms of Payment**

2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.

2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

*[Include the following clauses if Framework Agreement will be used:]*

2.3. For a single-year Framework Agreement, prices charged by the Supplier for Goods delivered and/or services performed under a Call-Off shall not vary from the prices quoted by the Supplier in its bid.

2.4. For multi-year Framework Agreement, prices charged by the Supplier for Goods delivered and/or services performed under a Call-Off shall not vary from the prices quoted by the Supplier during conduct of Mini-Competition.

### **3. Performance Security**

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184. *{[Include if Framework Agreement will be used:] In the case of Framework Agreement, the Bidder may opt to furnish the performance security or a Performance Securing Declaration as defined under the Guidelines on the Use of Framework Agreement.}*





**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

## OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)

### 4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project *{[Include if Framework Agreement will be used:]*or Framework Agreement} specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the SCC, **Section IV (Technical Specifications)** shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

### 5. Warranty

- 6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.
- 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

### 6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

***Section V. Special  
Conditions of  
Contract***



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

***Special Conditions of Contract***

GCC Clause	
1	<p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).</p> <p>For purposes of this Clause the Procuring Entity's Representative at the Project Site is <b>PGSO Warehouse</b>.</p> <p><b>Incidental Services –</b></p> <p>The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements:</p> <p><i>Select appropriate requirements and delete the rest.</i></p> <ul style="list-style-type: none"><li>a. performance or supervision of on-site assembly and/or start-up of the supplied Goods;</li><li>b. furnishing of tools required for assembly and/or maintenance of the supplied Goods;</li><li>c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;</li><li>d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and</li></ul> <p><i>Select appropriate requirements and delete the rest.</i></p> <ul style="list-style-type: none"><li>e. performance or supervision of on-site assembly and/or start-up of the supplied Goods;</li><li>f. furnishing of tools required for assembly and/or maintenance of the supplied Goods;</li><li>g. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;</li><li>h. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and</li></ul>

✓



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

- i. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.
- j. *[Specify additional incidental service requirements, as needed.]*

The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

**Spare Parts –**

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

*Select appropriate requirements and delete the rest.*

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
  - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
  - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI (Schedule of Requirements)** and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of *[indicate here the time period specified. If not used indicate a time period of three times the warranty period]*.

Spare parts or components shall be supplied as promptly as possible, but in any case, within *[insert appropriate time period]* months of placing the order.



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

	<p><b>Packaging –</b></p> <p>The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.</p> <p>The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.</p> <p>The outer packaging must be clearly marked on at least four (4) sides as follows:</p> <p>Name of the Procuring Entity Name of the Supplier Contract Description Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications</p>
	<p>A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.</p> <p><b>Transportation –</b></p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.</p> <p>Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.</p>



**Republic of the Philippines**  
**PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

	<p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p> <p><b>Intellectual Property Rights –</b></p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
2.2	The terms of payment shall be <b>Credit Basis.</b>
4	The inspections and tests that will be conducted are: <i>[Indicate the applicable inspections and tests]</i>

2



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

***Section VI. Schedule  
of Requirements***

**PGO - BIDS AND AWARDS COMMITTEE**

**Schedule of Requirements**

The delivery schedule as per IB NO.: G-245-22 30 Calendar days

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
PR# (PHO)					
1	Clinical Height and Weighing Scale  *Height rod in inches and centimeters *Platform cover is easily removed for cleaning *Die-cast beam can be easily read from either side *Max weight: 160 kg *Scale division (D): 100 g *Height range to be measured: 75-200cm *Minimum value of height per division: 0.5cm *Size of packaging box: 1500 x 510 x 155cm *Load platform area (LxW): 375 x 275 mm *Measurement(LxWxH): 530x275x1485mm *Net weight: 15kg *Gross weight: 17kg	4	unit		
2	CT Single Syringe Injection System, Pedestal Type with Console  <i>Overview Features:</i> *Touch Screen/Color LCD *Simple Operation *Friendly design *Easy to understand result and easy protocol entry *Improves throughput minimizes error *Easy to learn and less training  Console: *Flow rate: 0.1-10mL/sec *Volume: 1-200mL *Pressure Limit: 10-300 PSI *Rise Time: OFF: 0.5 sec Fixed; ON: 2 sec Fixed *Real Time Pressure Graph: YES *Protocol Memory: 160 Protocols for five (5) users *Scan Time: 0-300 sec *Remote Injection Start/Abort/Hold: YES *Synchronization Features: YES  Powerhead: *Air Check Button: YES *Injection Start/Abort/Hold: YES *Syringe Adapter: 100/200 mL *Contrast Warmer: YES  At least 1 year warranty on parts and services; Installation, training and decommissioning included; After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.	1	unit		
3	Hospital Emergency Cart Made of Stainless Steel HLW ( 105 cm x 25 cm x 36 cm ) 3 Drawers and 2 doors Stainless Top	4	unit		
4	Hot Plate with Magnetic Stirrer 220V *5" stainless steel with ceramic coated plate *Heating temperature range 50°C - 320°C *LED speed display *Brushless DC motor	1	unit		

2



**PGO - BIDS AND AWARDS COMMITTEE**

**Schedule of Requirements**

The delivery schedule as per IB NO.: G-245-22 30 Calendar days

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
	*Motor Rating Output: at least 20W *Voltage 200-240/650W *Speed range 50-1500 rpm  At least 1 year warranty on parts and services				
5	Mc-intosh Fiber Optic Laryngoscope Adult *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case	5	unit		
6	Mc-intosh Fiber Optic Laryngoscope Pediatric *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case	1	unit		
7	Uninterrupted Power Supply Tender for CT Scan  •High frequency online type UPS used in medical equipment industry •Compatible with existing 16 Slice CT Scan machine •Must be three-phase with neutral plus ground  *Capacity: at least 80KVA *4 hours back up time and 2 hours scanning  *Nominal Voltage Input: at least 380VAC phase to phase and 220VAC phase to neutral or ground *Voltage Range: Full load: at least 305VAC phase to phase and 176VAC phase to neutral or ground; De-rating: at least 208 VAC 70 percent phase to phase and 120VAC phase to neutral or ground *Input Voltage Comeback: Low loss volatge +10V, high loss -10V *Frequency Range: 40 to 70 Hertz *Input Synchronization Range: +/-1Hz,+/-2Hz,+/-4Hz and user programmable depends on the compatibility needs *Power Factor: 1.0 *THDi at 100 percent load: Must be less than 3 percent  *Nominal bypass voltage: at least 380VAC phase to phase and 220 phase to neutral or ground *Nominal bypass voltage range: +15% upper limit and -20% lower limit *Nominal output frequency: 60 Hz *Nominal output frequency range: +/-1Hz,+/-2Hz,+/-4Hz *Bypass synchronization range: +/-1Hz,+/-2Hz,+/-4Hz  *Output voltages: at least 380VAC phase to phase and 230VAC phase to neutral or ground *AC regulation range: Less than 1% at typical balance load and less than 2% for unbalance load *Output frequency range: 50-60Hz  *Frequency range on battery mode: 60Hz +/-1	1	unit		

**PGO - BIDS AND AWARDS COMMITTEE**

**Schedule of Requirements**

The delivery schedule as per IB NO.: G-245-22 30 Calendar days

Item No.	Item and Description	Qty	Unit	Delivered	
				Weeks	Months
	<p>*Overload AC mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload</p> <p>*Overload battery mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload</p> <p>*Current crest ratio: 3:1 max</p> <p>*Harmonic distortion: 100% linear load should be less than 2% and 100% non linear load should be less than 4%</p> <p>*Transfer time AC Line to battery: 0ms</p> <p>*Battery Mode backup time: Must not be less than 1hour CT scan scanning mode and 4 hours on CT scan standby mode</p> <p>*Battery type: VRLA *Battery number: at least 40 pieces with rated specs of 12VDC and 150AH</p> <p>*Battery Cabinet: Must be customizable depends on the allocated space for electrical room</p> <p>*Maximum charging current: 24A</p> <p>*Floating voltage: 2.29V/cell</p> <p>*Boost charging: 2.35V/cell</p> <p>*Charger voltage regulation: Less than 1% *Ripple voltage: Less than 1% *Ripple current: Less than 5</p> <p>*UPS controller cabinet dimension: Should not be more than 950cmX570cmX1020cm</p> <p>*UPS controller cabinet weight: Should not be more than 195kg</p> <p>*Environmental operating temperature: 0-40 degrees Celcius</p> <p>*Environmental operating humidity: Should be less than 95% and non-condensing</p> <p>*Operating altitude: Should be less than 1000meters</p> <p>*Acoustic noise level: Should be less than 70dB at 1 meter distance</p> <p>*UPS smart system management: RS232 or USB and should support Windows, Linux and IOS operating system</p> <p>*Power management: Must have and optional SNMP module</p> <p>3 years warranty (1 year parts and labor + 2 years unlimited support on equipment) 1 year warranty on batteries;                      Installation, training and decommissioning included;                      After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</p>				
8	<p>Vaccine Storage Refrigerator</p> <p>Specs:</p> <p>*Type: Upright, Double Glass Door</p> <p>*Capacity: At least 940 Liters</p> <p>*Temperature Range: 2 degrees Celcius to 8 degrees Celcius</p> <p>*Forced air cooling, auto defrost &amp; microprocessor control</p> <p>*Bright LED display</p> <p>*Dimensions (WxDxH): 1130 x 755 x 1980 mm or manufacturer's standard dimensions equivalent</p> <p>*Power: 208-230/60</p> <p>*Good for at least 9, 450 vials</p> <p>*With 1 unit of AVR 3KVA</p> <p>1 YEAR WARRANTY; PREVENTIVE MAINTENANCE (SEMI-ANNUAL); After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</p>	1	unit		
	*****				

N



**Republic of the Philippines  
PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

# *Section VII. Technical Specifications*

**PGO - BIDS AND AWARDS COMMITTEE**

**Technical Specifications**

IAEB NO.: \_\_\_\_\_ Date of Opening: May 31, 2022 Quotation: By Item

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	PR# (PHO)				
1	Clinical Height and Weighing Scale  *Height rod in inches and centimeters *Platform cover is easily removed for cleaning *Die-cast beam can be easily read from either side *Max weight: 160 kg *Scale division (D): 100 g *Height range to be measured: 75-200cm *Minimum value of height per division: 0.5cm *Size of packaging box: 1500 x 510 x 155cm *Load platform area (LxW): 375 x 275 mm *Measurement(LxWxH): 530x275x1485mm *Net weight: 15kg *Gross weight: 17kg		4	unit	
2	CT Single Syringe Injection System, Pedestal Type with Console  <i>Overview Features:</i> *Touch Screen/Color LCD *Simple Operation *Friendly design *Easy to understand result and easy protocol entry *Improves throughput minimizes error *Easy to learn and less training  Console: *Flow rate: 0.1-10mL/sec *Volume: 1-200mL *Pressure Limit: 10-300 PSI *Rise Time: OFF: 0.5 sec Fixed; ON: 2 sec Fixed *Real Time Pressure Graph: YES *Protocol Memory: 160 Protocols for five (5) users *Scan Time: 0-300 sec *Remote Injection Start/Abort/Hold: YES *Synchronization Features: YES  Powerhead: *Air Check Button: YES *Injection Start/Abort/Hold: YES *Syringe Adapter: 100/200 mL *Contrast Warmer: YES  <i>At least 1 year warranty on parts and services; Installation, training and decommissioning included; After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</i>		1	unit	
3	Hospital Emergency Cart <i>Made of Stainless Steel</i> HLW (105cm x 25 cm x 36 cm) 3 Drawers and 2 doors Stainless Top		4	unit	
4	Hot Plate with Magnetic Stirrer 220V *5" stainless steel with ceramic coated plate *Heating temperature range 50°C - 320°C		1	unit	

**PGO - BIDS AND AWARDS COMMITTEE**

**Technical Specifications**

IAEB NO.: \_\_\_\_\_ Date of Opening: May 31, 2022 Quotation: By Item

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	*LED speed display *Brushless DC motor *Motor Rating Output: at least 20W *Voltage 200-240/650W *Speed range 50-1500 rpm  At least 1 year warranty on parts and services				
5	Mc-intosh Fiber Optic Laryngoscope Adult *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case		5	unit	
6	Mc-intosh Fiber Optic Laryngoscope Pediatric *Halogen Illumination *Manufactured from AISI 304 Stainless-Steel *Halogen Illumination 2.5V *Bulb is inside the handle in order to simplify the sterilization of the blades. *Blades and handles are interchangeable with other leading manufacturer *With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case		1	unit	
7	Uninterrupted Power Supply Tender for CT Scan  •High frequency online type UPS used in medical equipment industry •Compatible with existing 16 Slice CT Scan machine •Must be three-phase with neutral plus ground  *Capacity: at least 80KVA *4 hours back up time and 2 hours scanning  *Nominal Voltage Input: at least 380VAC phase to phase and 220VAC phase to neutral or ground *Voltage Range: Full load: at least 305VAC phase to phase and 176VAC phase to neutral or ground; De-rating: at least 208 VAC 70 percent phase to phase and 120VAC phase to neutral or ground *Input Voltage Comeback: Low loss voltage +10V, high loss -10V *Frequency Range: 40 to 70 Hertz *Input Synchronization Range: +/-1Hz, +/-2Hz, +/-4Hz and user programmable depends on the compatibility needs *Power Factor: 1.0 *THDi at 100 percent load: Must be less than 3 percent  *Nominal bypass voltage: at least 380VAC phase to phase and 220 phase to neutral or ground *Nominal bypass voltage range: +15% upper limit and -20% lower limit		1	unit	

1

**PGO - BIDS AND AWARDS COMMITTEE**

**Technical Specifications**

AEB NO.:

Date of Opening: May 31, 2022

Quotation: By Item

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	<p>*Nominal output frequency: 60 Hz                      *Nominal output frequency range: +/-1Hz,+/-2Hz,+/-4Hz                      *Bypass synchronization range: +/-1Hz,+/-2Hz,+/-4Hz</p> <p>*Output voltages: at least 380VAC phase to phase and 230VAC phase to neutral or ground                      *AC regulation range: Less than 1% at typical balance load and less than 2% for unbalance load                      *Output frequency range: 50-60Hz</p> <p>*Frequency range on battery mode: 60Hz +/-1                      *Overload AC mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload                      *Overload battery mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload                      *Current crest ratio: 3:1 max                      *Harmonic distortion: 100% linear load should be less than 2% and 100% non linear load should be less than 4%</p> <p>*Transfer time AC Line to battery: 0ms                      *Battery Mode backup time: Must not be less than 1hour CT scan scanning mode and 4 hours on CT scan standby mode                      *Battery type: VRLA *Battery number: at least 40 pieces with rated specs of 12VDC and 150AH                      *Battery Cabinet: Must be customizable depends on the allocated space for electrical room                      *Maximum charging current: 24A                      *Floating voltage: 2.29V/cell                      *Boost charging: 2.35V/cell                      *Charger voltage regulation: Less than 1% *Ripple voltage: Less than 1% *Ripple current: Less than 5</p> <p>*UPS controller cabinet dimension: Should not be more than 950cmX570cmX1020cm                      *UPS controller cabinet weight: Should not be more than 195kg                      *Environmental operating temperature: 0-40 degrees Celcius                      *Environmental operating humidity: Should be less than 95% and non-condensing                      *Operating altitude: Should be less than 1000meters                      *Acoustic noise level: Should be less than 70dB at 1 meter distance                      *UPS smart system management: RS232 or USB and should support Windows, Linux and IOS operating system                      *Power management: Must have and optional SNMP module</p> <p>3 years warranty (1 year parts and labor + 2 years unlimited support on equipment) 1 year warranty on batteries; Installation, training and decommissioning included; After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</p>				
8	<p>Vaccine Storage Refrigerator                      Specs:                      *Type: Upright, Double Glass Door</p>		1	unit	

**PGO - BIDS AND AWARDS COMMITTEE**

**Technical Specifications**

AEB NO.:

Date of Opening: May 31, 2022

Quotation: By Item

Item No.	Item and Description	Brand Name (Mandatory)	Qty	Unit	Statement of Compliance
	*Capacity: At least 940 Liters *Temperature Range: 2 degrees Celcius to 8 degrees Celcius *Forced air cooling, auto defrost & microprocessor control *Bright LED display *Dimensions (WxDxH): 1130 x 755 x 1980 mm or manufacturer's standard dimensions equivalent *Power: 208-230/60 *Good for at least 9, 450 vials *With 1 unit of AVR 3KVA  1 YEAR WARRANTY; PREVENTIVE MAINTENANCE (SEMI-ANNUAL); After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.				
	*****				

P



**Republic of the Philippines**  
**PROVINCE OF DAVAO DEL SUR**

Room 4, Executive Bldg., Brgy. Matti, Digos City, Davao del Sur  
Mobile Nos. 0905-229-0526 (Globe) / 0908-332-2024 (Smart)  
Telefax: (082) 553-9579/ Website: [www.davaodelsur.gov.ph](http://www.davaodelsur.gov.ph)  
Email: [bac.davaodelsur2@gmail.com](mailto:bac.davaodelsur2@gmail.com)

**OFFICE OF THE BIDS AND AWARDS COMMITTEE (BAC)**

***Section VIII.***  
***Checklist of Technical***  
***and Financial***  
***Documents***



**Price Schedule for Goods Offered from Within the Philippines**  
*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

**For Goods Offered from Within the Philippines**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x
1	<b>PR# (PHO)</b>  <b>Clinical Height and Weighing Scale</b>  *Height rod in inches and centimeters *Platform cover is easily removed for cleaning *Die-cast beam can be easily read from either side *Max weight: 160 kg *Scale division (D): 100 g *Height range to be measured: 75-200cm *Minimum value of height per division: 0.5cm *Size of packaging box: 1500 x 510 x 155cm *Load platform area (LxW): 375 x 275 mm *Measurement(LxWxH): 530x275x1485mm *Net weight: 15kg *Gross weight: 17kg		4						
2	<b>CT Single Syringe Injection System, Pedestal Type with Console</b>  <i>Overview Features:</i> *Touch Screen/Color LCD *Simple Operation *Friendly design *Easy to understand result and easy protocol entry *Improves throughput minimizes error *Easy to learn and less training  <i>Console:</i> *Flow rate: 0.1-10mL/sec *Volume: 1-200mL *Pressure Limit: 10-300 PSI *Rise Time: OFF: 0.5 sec Fixed; ON: 2 sec Fixed *Real Time Pressure Graph: YES		1						

**Price Schedule for Goods Offered from Within the Philippines**  
*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

For Goods Offered from Within the Philippines

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
	*Protocol Memory: 160 Protocols for five (5) users *Scan Time: 0-300 sec *Remote Injection Start/Abort/Hold: YES *Synchronization Features: YES  Powerhead: *Air Check Button: YES *Injection Start/Abort/Hold: YES *Syringe Adapter: 100/200 mL *Contrast Warmer: YES  <b>At least 1 year warranty on parts and services;            Installation, training and decommissioning included;            After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</b>								
3	<b>Hospital Emergency Cart</b> Made of Stainless Steel HLW ( 105 cm x 25 cm x 36 cm ) 3 Drawers and 2 doors Stainless Top IV hook with Bumper		4						
4	<b>Hot Plate with Magnetic Stirrer 220V</b> *5" stainless steel with ceramic coated plate *Heating temperature range 50°C - 320°C *LED speed display *Brushless DC motor *Motor Rating Output: at least 20W *Voltage 200-240/650W *Speed range 50-1500 rpm		1						

**Price Schedule for Goods Offered from Within the Philippines**  
*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

For Goods Offered from Within the Philippines

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)
	<b>At least 1 year warranty on parts and services</b>								
5	<b>Mc-intosh Fiber Optic Laryngoscope Adult</b> <i>*Halogen Illumination</i> <i>*Manufactured from AISI 304 Stainless-Steel</i> <i>*Halogen Illumination 2.5V</i> <i>*Bulb is inside the handle in order to simplify the sterilization of the blades.</i> <i>*Blades and handles are interchangeable with other leading manufacturer</i> <i>*With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case</i>		5						
6	<b>Mc-intosh Fiber Optic Laryngoscope Pediatric</b> <i>*Halogen Illumination</i> <i>*Manufactured from AISI 304 Stainless-Steel</i> <i>*Halogen Illumination 2.5V</i> <i>*Bulb is inside the handle in order to simplify the sterilization of the blades.</i> <i>*Blades and handles are interchangeable with other leading manufacturer</i> <i>*With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case</i>		1						
7	<b>Uninterrupted Power Supply Tender for CT Scan</b> <i>*High frequency online type UPS used in medical equipment industry</i> <i>*Compatible with existing 16 Slice CT Scan machine</i> <i>*Must be three-phase with neutral plus ground</i>		1						

**Price Schedule for Goods Offered from Within the Philippines**  
*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

**For Goods Offered from Within the Philippines**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
	<p>*Capacity: at least 80KVA                      *4 hours back up time and 2 hours scanning</p> <p>*Nominal Voltage Input: at least 380VAC phase to phase and 220VAC phase to neutral or ground                      *Voltage Range: Full load: at least 305VAC phase to phase and 176VAC phase to neutral or ground; De-rating: at least 208 VAC 70 percent phase to phase and 120VAC phase to neutral or ground                      *Input Voltage Comeback: Low loss volatige +10V, high loss -10V                      *Frequency Range: 40 to 70 Hertz                      *Input Synchronization Range: +/-1Hz,+/-2Hz,+/-4Hz and user programmable depends on the compatibility needs                      *Power Factor: 1.0                      *THDI at 100 percent load: Must be less than 3 percent</p> <p>*Nominal bypass voltage: at least 380VAC phase to phase and 220 phase to neutral or ground                      *Nominal bypass voltage range: +15% upper limit and -20% lower limit                      *Nominal output frequency: 60 HZ                      *Nominal output frequency range: +/-1Hz,+/-2Hz,+/-4Hz                      *Bypass synchronization range: +/-1Hz,+/-2Hz,+/-4Hz</p> <p>*Output voltages: at least 380VAC phase to phase and 230VAC phase to neutral or ground                      *AC regulation range: Less than 1% at typical balance load and less than 2% for unbalance load                      *Output frequency range: 50-60Hz</p> <p>*Frequency range on battery mode: 60Hz +/-1                      *Overload AC mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload</p>								

## Price Schedule for Goods Offered from Within the Philippines

*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

### For Goods Offered from Within the Philippines

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
	<p>*Overload battery mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload</p> <p>*Current crest ratio: 3:1 max</p> <p>*Harmonic distortion: 100% linear load should be less than 2% and 100% non linear load should be less than 4%</p> <p>*Transfer time AC Line to battery: 0ms</p> <p>*Battery Mode backup time: Must not be less than 1hour CT scan scanning mode and 4 hours on CT scan standby mode</p> <p>*Battery type: VRLA *Battery number: at least 40 pieces with rated specs of 12VDC and 150AH</p> <p>*Battery Cabinet: Must be customizable depends on the allocated space for electrical room</p> <p>*Maximum charging current: 24A</p> <p>*Floating voltage: 2.29V/cell</p> <p>*Boost charging: 2.35V/cell</p> <p>*Charger voltage regulation: Less than 1% *Ripple voltage: Less than 1% *Ripple current: Less than 5</p> <p>*UPS controller cabinet dimension: Should not be more than 950cmX570cmX1020cm</p> <p>*UPS controller cabinet weight: Should not be more than 195kg</p> <p>*Environmental operating temperature: 0-40 degrees Celsius</p> <p>*Environmental operating humidity: Should be less than 95% and non-condensing</p> <p>*Operating altitude: Should be less than 1000meters</p> <p>*Acoustic noise level: Should be less than 70dB at 1 meter distance</p> <p>*UPS smart system management: RS232 or USB and should support Windows, Linux and IOS operating system</p> <p>*Power management: Must have and optional SNMP module</p>								

## Price Schedule for Goods Offered from Within the Philippines

*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

**For Goods Offered from Within the Philippines**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 9) x (col 4)
8	<b>Vaccine Storage Refrigerator</b>  Specs: *Type: Upright, Double Glass Door *Capacity: At least 940 Liters *Temperature Range: 2 degrees Celcius to 8 degrees Celcius *Forced air cooling, auto defrost & microprocessor control *Bright LED display *Dimensions (WxDxH): 1130 x 755 x 1980 mm or manufacturer's standard dimensions equivalent *Power: 208-230/60 *Good for at least 9, 450 vials *With 1 unit of AVR 3KVA  <b>1 YEAR WARRANTY; PREVENTIVE MAINTENANCE (SEMI-ANNUAL): After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</b>		1						

**Price Schedule for Goods Offered from Within the Philippines**  
*[shall be submitted with the Bid if bidder is offering goods from within the Philippines]*

**For Goods Offered from Within the Philippines**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	10
Item	Description	Country of origin	Quantity	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Total Price, per unit (col 5+6+7+8)	Total Price delivered Final Destination (col 4) x (col 9)
	Charges: GENERAL FUND 2022 CAPITAL OUTLAY - MEDICAL EQUIPMENT 1-07-05-110 Purpose: FOR DSPH IMAGING DEPARTMENT, PHARMACY DEPARTMENT, LABORATORY DEPARTMENT AND NURSING SERVICES USE.								

Name: \_\_\_\_\_ Legal  
 Capacity: \_\_\_\_\_  
 Signature: \_\_\_\_\_ Duly  
 authorized to sign the Bid for and behalf of: \_\_\_\_\_

**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9	
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)	
1	<b>Clinical Height and Weighing Scale</b> *Height rod in inches and centimeters *Platform cover is easily removed for cleaning *Die-cast beam can be easily read from either side *Max weight: 160 kg *Scale division (D): 100 g *Height range to be measured: 75-200cm *Minimum value of height per division: 0.5cm *Size of packaging box: 1500 x 510 x 155cm *Load platform area (LxW): 375 x 275 mm *Measurement(LxWxH): 530x275x1485mm *Net weight: 15kg *Gross weight: 17kg			4					
2	<b>CT Single Syringe Injection System, Pedestal Type with Console</b> Overview Features: *Touch Screen/Color LCD *Simple Operation *Friendly design *Easy to understand result and easy protocol entry *Improves throughput minimizes error *Easy to learn and less training Console: *Flow rate: 0.1-10ml/sec *Volume: 1-200ml *Pressure Limit: 40-999 PSI		1						



**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)
	<p>*Rise Time: OFF: 0.5 sec: Fixed: ON: 2 sec: Fixed                      *Real Time Pressure Graph: YES                      *Protocol Memory: 160 Protocols for five (5) users                      *Scan Time: 0-300 sec                      *Remote Injection Start/Abort/Hold: YES                      *Synchronization Features: YES</p> <p>Powerhead:                      *Air Check Button: YES                      *Injection Start/Abort/Hold: YES                      *Syringe Adapter: 100/200 mL                      *Contrast Warmer: YES</p> <p><b>At least 1 year warranty on parts and services; Installation, training and decommissioning included; After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</b></p>							
3	<p><b>Hospital Emergency Cart</b>                      Made of Stainless Steel                      HLW ( 105 cm x 25 cm x 36 cm )                      3 Drawers and 2 doors                      Stainless Top                      IV hook with Bumper</p>		4					
4	<p><b>Hot Plate with Magnetic Stirrer 220V</b>                      *5" stainless steel with ceramic coated plate                      *Heating temperature range 50°C - 320°C                      *LED speed display                      *Brushless DC motor</p>		1					

**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)
	<p>*Motor Rating Output: at least 20W                      *Voltage 200-240/650W                      *Speed range 50-1500 rpm  <b>At least 1 year warranty on parts and services</b></p>							
5	<p><b>Mc-intosh Fiber Optic Laryngoscope Adult</b>                      *Halogen Illumination                      *Manufactured from AISI 304 Stainless-Steel                      *Halogen Illumination 2.5V                      *Bulb is inside the handle in order to simplify the sterilization of the blades.                      *Blades and handles are interchangeable with other leading manufacturer                      *With set of 1 (working length of 70mm), 2 (working length of 78mm), 3 (working length of 108mm), and 4 (working length of 133mm) blades supplied in green plastic case</p>		5					
6	<p><b>Mc-intosh Fiber Optic Laryngoscope Pediatric</b>                      *Halogen Illumination                      *Manufactured from AISI 304 Stainless-Steel                      *Halogen Illumination 2.5V                      *Bulb is inside the handle in order to simplify the sterilization of the blades.                      *Blades and handles are interchangeable with other leading manufacturer                      *With set of 0 (working length of 55mm), 1 (working length of 70mm), and 2 (working length of 78mm) blades supplied in green plastic case</p>		1					
7	<p><b>Uninterrupted Power Supply Tender for CT Scan</b></p>		1					

**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)
	<p><i>*High frequency online type UPS used in medical equipment industry</i></p> <p><i>*Compatible with existing 16 Slice CT Scan machine</i></p> <p><i>*Must be three-phase with neutral plus ground</i></p> <p><i>*Capacity: at least 80KVA</i></p> <p><i>*4 hours back up time and 2 hours scanning</i></p> <p><i>*Nominal Voltage Input: at least 380VAC phase to phase and 220VAC phase to neutral or ground</i></p> <p><i>*Voltage Range: Full load: at least 305VAC phase to phase and 176VAC phase to neutral or ground; De-rating: at least 208 VAC 70 percent phase to phase and 120VAC phase to neutral or ground</i></p> <p><i>*Input Voltage Comeback: Low loss volatge +10V, high loss -10V</i></p> <p><i>*Frequency Range: 40 to 70 Hertz</i></p> <p><i>*Input Synchronization Range: +/-1Hz,+/-2Hz,+/-4Hz and user programmable depends on the compatibility needs</i></p> <p><i>*Power Factor: 1.0</i></p> <p><i>*THDI at 100 percent load: Must be less than 3 percent</i></p> <p><i>*Nominal bypass voltage: at least 380VAC phase to phase and 220 phase to neutral or ground</i></p> <p><i>*Nominal bypass voltage range: +15% upper limit and -20% lower limit</i></p> <p><i>*Nominal output frequency: 60 Hz</i></p> <p><i>*Nominal output frequency range: +/-1Hz,+/-2Hz,+/-4Hz</i></p> <p><i>*Bypass synchronization range: +/-1Hz,+/-2Hz,+/-4Hz</i></p> <p><i>*Output voltages: at least 380VAC phase to phase and 230VAC phase to neutral or ground</i></p>							

**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)
	<p>*AC regulation range: Less than 1% at typical balance load and less than 2% for unbalance load                      *Output frequency range: 50-60Hz                      *Frequency range on battery mode: 60Hz +/-1                      *Overload AC mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload                      *Overload battery mode: 1hour for 105-110%, 10mins for 110-125%, 1min for 126-150% and 200ms for more than 150% overload                      *Current crest ratio: 3:1 max                      *Harmonic distortion: 100% linear load should be less than 2% and 100% non linear load should be less than 4%                      *Transfer time AC Line to battery: 0ms                      *Battery Mode backup time: Must not be less than 1hour CT scan scanning mode and 4 hours on CT scan standby mode                      *Battery type: VRLA *Battery number: at least 40 pieces with rated specs of 12VDC and 150AH                      *Battery Cabinet: Must be customizable depends on the allocated space for electrical room                      *Maximum charging current: 24A                      *Floating voltage: 2.29V/cell                      *Boost charging: 2.35V/cell                      *Charger voltage regulation: Less than 1% *Ripple voltage: Less than 1% *Ripple current: Less than 5                      *UPS controller cabinet dimension: Should not be more than 950cmX570cmX1020cm                      *UPS controller cabinet weight: Should not be more than 195kg</p>							

**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price: Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)
8	<p><b>Vaccine Storage Refrigerator</b></p> <p>Specs:                      *Type: Upright, Double Glass Door                      *Capacity: At least 940 Liters                      *Temperature Range: 2 degrees Celsius to 8 degrees Celsius                      *Forced air cooling, auto defrost &amp; microprocessor control                      *Bright LED display                      *Dimensions (WxDxH): 1130 x 755 x 1980 mm or manufacturer's standard dimensions equivalent                      *Power: 208-230/60                      *Good for at least 9, 450 vials</p>		1					
	<p>*Environmental operating temperature: 0-40 degrees Celsius                      *Environmental operating humidity: Should be less than 95% and non-condensing                      *Operating altitude: Should be less than 1000meters                      *Acoustic noise level: Should be less than 70dB at 1 meter distance                      *UPS smart system management: RS232 or USB and should support Windows, Linux and IOS operating system                      *Power management: Must have and optional SNMP module</p> <p><b>3 years warranty (1 year parts and labor + 2 years unlimited support on equipment) 1 year warranty on batteries;</b>                      Installation, training and decommissioning included;                      After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</p>							

**Price Schedule for Goods Offered from Abroad**  
*[shall be submitted with the Bid if bidder is offering goods from Abroad]*

. 2

**For Goods Offered from Abroad**

Name of Bidder \_\_\_\_\_ Project ID No. \_\_\_\_\_ Page \_\_\_\_\_ of \_\_\_\_\_

1	2	3	4	5	6	7	8	9
Item	Description	Country of origin	Quantity	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 4 x 5)	Unit Price Delivered Duty Unpaid (DDU)	Unit price: Delivered Duty Paid (DDP)	Total Price delivered DDP (col 4 x 8)
	*With 1 unit of AVR 3KVA <b>1 YEAR WARRANTY; PREVENTIVE MAINTENANCE (SEMI-ANNUAL); After sales service 24/7 and on-call if problem arises as needed; has services center within Region XI or General Santos City; Delivery Period: 30 calendar days.</b>							
	Charges: GENERAL FUND 2022 CAPITAL OUTLAY - MEDICAL EQUIPMENT 1-07-05-110 Purpose: FOR DSPH IMAGING DEPARTMENT, PHARMACY DEPARTMENT, LABORATORY DEPARTMENT AND NURSING SERVICES USE.							

Name: \_\_\_\_\_ Legal  
 Capacity: \_\_\_\_\_  
 Signature: \_\_\_\_\_ Duly  
 authorized to sign the Bid for and behalf of: \_\_\_\_\_